



Designation: D 6512 – 00

Standard Practice for Interlaboratory Quantitation Estimate¹

This standard is issued under the fixed designation D 6512; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice establishes a uniform standard for computing the interlaboratory quantitation estimate associated with $Z\%$ relative standard deviation (referred to herein as $IQE_{Z\%}$), and provides guidance concerning the appropriate use and application.

1.2 $IQE_{Z\%}$ is computed to be the lowest concentration for which a single measurement from a laboratory selected from the population of qualified laboratories represented in an interlaboratory study will have an estimated $Z\%$ relative standard deviation ($Z\%$ RSD, based on interlaboratory standard deviation), where Z is typically an integer multiple of 10, such as 10, 20, or 30, but Z can be less than 10. The $IQE_{10\%}$ is consistent with the quantitation approaches of Currie **(1)**² and Oppenheimer, et al **(2)**.

1.3 The fundamental assumption of the collaborative study is that the media tested, the concentrations tested, and the protocol followed in the study provide a representative and fair evaluation of the scope and applicability of the test method as written. Properly applied, the IQE procedure ensures that the IQE has the following properties:

1.3.1 *Routinely Achievable IQE Value*—Most laboratories are able to attain the IQE quantitation performance in routine analyses, using a standard measurement system, at reasonable cost. This property is needed for a quantitation limit to be feasible in practical situations. Representative laboratories must be included in the data to calculate the IQE.

1.3.2 *Accounting for Routine Sources of Error*—The IQE should realistically include sources of bias and variation that are common to the measurement process. These sources include, but are not limited to: intrinsic instrument noise, some “typical” amount of carryover error; plus differences in laboratories, analysts, sample preparation, and instruments.

1.3.3 *Avoidable Sources of Error Excluded*—The IQE should realistically exclude avoidable sources of bias and variation; that is, those sources that can reasonably be avoided

in routine field measurements. Avoidable sources would include, but are not limited to: modifications to the sample; modifications to the measurement procedure; modifications to the measurement equipment of the validated method, and gross and easily discernible transcription errors, provided there was a way to detect and either correct or eliminate them.

1.4 The IQE applies to measurement methods for which calibration error is minor relative to other sources, such as when the dominant source of variation is one of the following:

1.4.1 *Sample Preparation*, and calibration standards do not have to go through sample preparation.

1.4.2 *Differences in Analysts*, and analysts have little opportunity to affect calibration results (as is the case with automated calibration).

1.4.3 *Differences in Laboratories* (for whatever reasons), perhaps difficult to identify and eliminate.

1.4.4 *Differences in Instruments* (measurement equipment), such as differences in manufacturer, model, hardware, electronics, sampling rate, chemical processing rate, integration time, software algorithms, internal signal processing and thresholds, effective sample volume, and contamination level.

1.5 *Data Quality Objectives*—Typically, one would compute the lowest $\%$ RSD possible for any given dataset for a particular method. Thus, if possible, $IQE_{10\%}$ would be computed. If the data indicated that the method was too noisy, one might have to compute instead $IQE_{20\%}$, or possibly $IQE_{30\%}$. In any case, an IQE with a higher $\%$ RSD level (such as $IQE_{50\%}$) would not be considered, though an IQE with RSD $<10\%$ (such as $IQE_{1\%}$) would be acceptable. The appropriate level of $\%$ RSD may depend on the intended use of the IQE.

2. Referenced Documents

2.1 ASTM Standards:

D 2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D-19 on Water³

D 6091 Practice for 99 %/95 % Interlaboratory Detection Estimate (IDE) for Analytical Methods with Negligible Calibration Error³

E 1763 Guide for Interpretation and Use of Results from

¹ This practice is under the jurisdiction of ASTM Committee D19 on Water and is the direct responsibility of Subcommittee D19.02 on General Specifications, Technical Resources, and Statistical Methods.

Current edition approved Feb. 10, 2000. Published May 2000.

² The boldface numbers in parentheses refer to the list of references at the end of this standard.

³ *Annual Book of ASTM Standards*, Vol 11.01.

Interlaboratory Testing of Chemical Analysis Methods⁴

3. Terminology

3.1 *Z % Interlaboratory Quantitation Estimate (IQE_{Z %})*, also denoted “LQ,” for “Limit of Quantitation” in accordance with Currie (1)—The lowest concentration for which a single measurement from a laboratory selected from the population of qualified laboratories represented in an interlaboratory study will have an estimated *Z %* relative standard deviation (*Z %* RSD, based on interlaboratory standard deviation).

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *Censored Measurement*—A measurement that is not reported numerically nor is reported missing, but is stated as a nondetect or a less-than (for example, “less than 0.1 ppb”). There are two reasons why the measurement may not be reported numerically. Either the measurement was considered insufficiently precise or accurate (these kinds of data should not be censored), or the identification of the analyte was suspect (these kinds of data should be censored). See §6.2.3.1. A reported “less than” may have the same meaning as a non-reported measurement, but a reported “less than” also implies (perhaps erroneously) that any concentration greater than or equal to the accompanying value (for example, 0.1 ppb) can be measured, and will be reported numerically.

3.2.2 *Quantitation Limit (QL) or Limit of Quantitation (LQ)*—A numerical value, expressed in physical units or proportion, intended to represent the lowest level of reliable quantitation. The IQE is an example of a QL.

4. Summary of Practice

4.1 Every ASTM Committee D-19 test method is evaluated to determine precision and bias by conducting a collaborative study, in accordance with Practice D 2777. That study, or a similar collaborative study, can also be used to evaluate the lowest concentration level of reliable quantitation for a test method, referred to herein as the interlaboratory quantitation estimate (IQE). Such a study must include concentrations suitable for modeling the uncertainty of mean recovery of interlaboratory measurement, preferably without extrapolation. The study must also be planned and conducted to allow the known, routine sources of measurement variability to be observed at typical levels of influence. After the study is conducted, outlying laboratories and individual measurements should be eliminated, using an accepted, scientifically based procedure for outlier removal, such as found in Practice D 2777. The IQE computations must be based on retained data from at least six independent laboratories at each concentration level.

4.2 Retained data are analyzed to identify and fit one of three proposed interlaboratory standard deviation (ILSD) models. These models describe the relationship between the interlaboratory standard deviation of measurements and the true concentration, *T*. The identification process involves evaluating the models in order, from simplest to most complex: constant,

straight-line, and hybrid (proposed by Rocke and Lorenzato (3)). Evaluation includes statistical significance and residual analysis.

4.3 The chosen model is used to predict the standard deviation of interlaboratory measurements at any true concentration within the study concentration range. If interlaboratory standard deviations change systematically with respect to the true concentration (that is, they are NOT constant), the predictions are used to generate weights for fitting the mean-recovery relationship (the assumed straight-line relationship between measured concentration and true concentration), using weighted least squares. (Otherwise, ordinary least squares is used.) The mean-recovery curve is evaluated for statistical significance, for lack of fit, and for residual patterns. The ILSD model is also used to estimate the interlaboratory standard deviation at concentrations within the concentration range. Either a direct or interactive algorithm (depending on the model) is used to compute IQE_{10 %}, the lowest concentration with estimated RSD = 10 % (*Z* = 10). If there is no such concentration, then IQE_{20 %} is computed instead, or IQE_{30 %}, if necessary. If supported by the data quality objectives (DQOs), IQE_{Z %} may be computed for some *Z* < 10.

5. Significance and Use

5.1 Appropriate application of this practice should result in an IQE achievable by most laboratories properly using the test method studied. That is, most laboratories should be capable of measuring concentrations greater than IQE_{Z %} with RSD = *Z %* or less. The IQE provides the basis for any prospective use of the test method by qualified laboratories for reliable quantitation of low-level concentrations of the same analyte as the one studied in this practice, and same media (matrix).

5.2 The IQE values may be used to compare the quantitation capability of different methods for analysis of the same analyte in the same matrix. The IQE is not an indicator of individual laboratory performance.

5.3 The IQE procedure should be used to establish the interlaboratory quantitation capability for any application of a method where interlaboratory quantitation is important to data use. The intent of the IQE is not to set reporting limits.

6. Procedure

6.1 The following procedure has stages described in the following paragraphs: 6.2—IQE Study Plan, Design, and Protocol; 6.3—Conduct the IQE Study, Screen the Data, and Choose a Model; and 6.4—Compute the IQE. A flowchart of the procedure is shown in Fig. 1.

6.2 *IQE Study Plan, Design, and Protocol:*

6.2.1 *Choose Analyte, Matrix, and Method*—At least one analyte of interest is selected, typically one for which there is interest in trace or near-trace levels of concentration, such as toxic materials that are controlled and regulated. For each analyte, an approximate maximum true concentration is selected, based on these considerations:

6.2.1.1 The anticipated IQE should be exceeded by a factor of 2 or more,

6.2.1.2 A single model, (ideally a straight-line model in true concentration, *T*) should describe mean recovery (that is, mean

⁴ Annual Book of ASTM Standards, Vol 03.06.




FIG. 1 Flowchart of IQE Procedure

measured concentration) for the entire range of concentrations, from zero to the selected maximum concentration.

NOTE 1—The IQE procedure uses the straight-line model for mean recovery, thus implicitly assuming that a straight line is adequate. Thus, the IQE would not be appropriate for cases where this assumption is unreasonable. For example, it would not hold for cases where there was systematic bias for most or all laboratories, such as a tendency to report values that are too high for some portion of the concentration range.

6.2.1.3 A single model in true concentration should describe the standard deviation of interlaboratory measurements for the entire range of concentrations, from zero to the selected maximum concentration.

6.2.1.4 The concentration range must be sufficient to enable statistically significant coefficients to be estimated for the ILSD model and mean-recovery model. At least one matrix of interest is also selected, and an accepted standard analytical method for those analytes is selected for study. If there is no possibility of matrix interference, then it may only be necessary to determine a list of acceptable matrices that can be used, instead of selecting a specific matrix. For example, for a particular analyte, concentration range, and method, it may be supposed that reagent waters from different laboratories are indistinguishable. However, that assumption may not hold for another analyte or another concentration range.

6.2.2 *Choose IQE Study Design*—The design should be based (if possible) on an anticipated ILSD model. Section 7 of Practice D 2777 can be followed for the study design and protocol. The anticipated form of the ILSD model (the relationship between interlaboratory measurement standard deviation and true concentration) can help in choosing an IQE study

design. Three models are proposed herein for the relationship between the interlaboratory standard deviation of measurements and the true concentration: constant, straight-line (increasing), and hybrid (increasing). See 6.3.3 for details. Chemistry, physics, empirical evidence, or informed judgment may make one model more plausible than others. However, it may not be possible to anticipate the relationship between standard deviation and true concentration.

6.2.2.1 Select an IQE study design that has enough distinct concentration levels to assess statistical lack of fit of the models (see Draper and Smith (4)). Recommended designs are: (a) the “semi-geometric” design at five or more true concentrations, $\{T_1, T_2, \text{ and so forth}\}$, such as: $\{0, IQE_0/4, IQE_0/2, IQE_0, 2 \times IQE_0, 4 \times IQE_0, 8 \times IQE_0\}$, where IQE_0 is an initial estimate of the IQE (such as $10s'$ where s' is the interlaboratory measurement standard deviation at a trace-level, nonzero concentration); (b) equi-spaced design: $\{0, IQE_0/2, IQE_0, (3/2) \times IQE_0, 2 \times IQE_0, (5/2) \times IQE_0, 3 \times IQE_0\}$; and (c) any other design with at least five concentrations, provided that the design includes at least one concentration approximately equal to $2 \times IQE_0$, at least one nonzero concentration below IQE_0 , and one blank, or unspiked sample. Preferably, the design will have at least seven concentrations, including a blank.

6.2.2.2 The study's concentration levels must either be known (true concentration levels), or knowable, after the fact. A concentration is considered known if reference standards can be purchased or constructed, and knowable if an accurate determination can be made (for example, the median value from many laboratories, or results from a recognized laboratory, such as NIST, using a high-accuracy method).

6.2.3 *Choose Protocol*—The protocol should follow Section 7 of Practice D 2777. The protocol should include design run order and details on when the system is to be purged, have extra blanks run, and so on. It should take into consideration possible problems with carryover, study cost (in time and money), and the time constants of drift of the measurement system or degradation of the sample.

6.2.3.1 For purposes of the collaborative study, the study supervisor should provide instructions to participating laboratories to disable (if possible) any internal reporting limits or any other data-censoring thresholds (such as an “instrument detection limit”) that are used to determine whether a numerical measurement is to be reported as a number, or as a nondetect or less-than (that is, the number is censored). If censoring is unavoidable, the laboratory censoring threshold must be reported with the study data. However, qualitative criteria used by the method to identify and discriminate among analytes are separate criteria, and must be satisfied in accordance with the method.

6.2.4 *Choose Allowable Sources of Variation*—It is assumed that, collectively, the many sources of variation will cause interlaboratory measurements at any true concentration to be Normally distributed. The number of laboratories providing usable data must be maximized in order for the study to capture representative between-laboratory variation adequately. Ordinary within-laboratory variation must be allowed to affect the measurement process, as happens in routine measurement. Ideally, there would be many laboratories, and each measurement at each laboratory would be made as a routine measurement, made by a different analyst using a different (qualified) measurement system on a different day, in random order, without the analyst being aware of the true value, or even that the sample was part of a special study.

6.2.4.1 As emphasized in Practice D 2777, maximizing the number of participating laboratories is often the most important thing that can be done to guarantee a successful study. The number of laboratories providing a full set of usable data will typically fall short of the number of participating laboratories. A minimum of ten participating laboratories is recommended.

6.2.4.2 To the extent possible, the study should be conducted so as to mimic routine laboratory measurement, particularly if the method is labor-intensive, as opposed to a highly automated method. That is, not only should the analysts not be aware of the true concentrations of these samples, but also they should not know that they are measuring special, study samples. These restrictions minimize the risk of extra-care distortion of data so common in analytical studies. However, it is recommended that the participating analysts be told to disable data-censoring limits, because there may or may not be some low concentrations in the study samples (see 6.2.3.1).

6.2.4.3 For each laboratory, the maximum possible number of qualified analysts should be involved in the study, since there are variations that may be allowed by the method, may be practiced by different analysts, and will be seen in routine analyses.

6.2.4.4 For each laboratory, the maximum possible number of qualified measurement systems should be used, since there

are model-to-model and instrument-to-instrument differences in equipment and maintenance, as will be seen in routine analyses.

6.2.4.5 For each laboratory, the IQE study should be scheduled to span the maximum possible number of days consistent with holding-time constraints, since day-to-day changes in analytical laboratory environmental conditions, contamination, solvent purity, and other factors can affect measurements, and will be seen in routine analyses.

6.3 *Conduct the IQE Study, Screen the Data, and Choose a Model:*

6.3.1 The IQE study should be conducted in accordance with Section 9 of Practice D 2777. Blank correction should not be performed by the laboratories, unless the method requires this subtraction in order to perform the test. Each laboratory should supply method-blank data along with the uncorrected measurement values, and the study supervisor can determine whether the reported measurements should be corrected.

6.3.2 The IQE study data should be screened in accordance with the initial subsections (relating to removing data) of Section 10 of Practice D 2777. (Proceed to Section 6.5 of the IQE Practice if, for any concentration, more than 10 % of the retained measurements are nondetects or less-thans.)

6.3.3 *Identify and Fit the ILSD Model*—The ILSD model should be identified and its coefficients should be estimated by using the following procedure. See Draper and Smith (4) and Caulett and Boddy (5) for more discussion of how to model standard deviations and how to do weighted least squares (WLS) in analytical chemistry. See Carroll and Ruppert (6) for further discussion of standard-deviation modeling. The ILSD model is an attempt to characterize the unknown (or partly known) relationship ($\sigma = G(T)$) between the actual standard deviation of interlaboratory measurement and true concentration. The model is used for two purposes: to provide weights for the WLS regression to fit the mean-recovery model, and to provide the interlaboratory standard-deviation estimates required to determine the IQE.

6.3.3.1 Three ILSD models are proposed. The identification process considers (that is, fits then evaluates) each model in turn, from simplest to most complex, until a suitable model is found. Prior knowledge can be combined with empirical results to influence the selection of a model if a suitable referenced publication can be cited. The model order is as follows:

(a) *Constant Model for the ILSD (Model A):*

$$s = g + \text{error} \quad (1)$$

where:

s = the sample standard deviation for interlaboratory measurements,

g = estimated constant, and

“error” is included for arithmetic completeness, since the model will not hold exactly. Interlaboratory standard deviation does not change with concentration, resulting in a relative standard deviation that declines with increasing T

(b) *Straight-line Model for the ILSD (Model B):*

$$s = g + hT + \text{error} \quad (2)$$

where: g and h = fitted constants.

Interlaboratory standard deviation increases linearly with concentration, resulting in an asymptotically constant relative standard deviation as T increases.

(c) *Hybrid Model for the ILSD (Model C):*

$$s = (g^2 + [hT]^2)^{1/2} + \text{error} \quad (3)$$

where the positive square root is taken; g and h are fitted constants. Interlaboratory standard deviation increases with concentration, at first slowly, then achieving proportional increase. This behavior also results in a relative standard deviation that initially declines as the concentration increases from zero, then asymptotically approaches a constant level. The Hybrid Model, the form of which was developed by Roche and Lorenzato (3) is so-named because it incorporates two things: additive error with constant standard deviation (coefficient g), and multiplicative error with increasing standard deviation (coefficient h).

NOTE 2—The Hybrid Model used the form of Roche and Lorenzato, but not necessarily the same assumptions for error distribution. The Hybrid Model is also the same as the General Analytical Error Model of Guide E 1763

In all cases, it is assumed that $g > 0$ (though this constraint is irrelevant for the Hybrid Model). A value of $g < 0$ has no practical interpretation, and may indicate that a different ILSD model should be used. Furthermore, it is assumed that g is not underestimated because of censored data among measurements of blanks or other low-concentration samples. (Censoring is addressed in 6.2.3.1, 6.3.2, and 6.5).

If $h < 0$, then it must be significantly less than zero (statistically), in which case the Constant Model (Model A) should be evaluated.

6.3.3.2 ILSD-Model Identification and Fitting Procedure:

See Section 10 for a detailed example, using the Hybrid Model for the ILSD.

(a) Merge all retained IQE study data (after possible elimination of some data in accordance with 6.3.2).

(b) For each true concentration, T , compute the adjusted interlaboratory sample standard deviation, s_k , an estimate of the true underlying interlaboratory measurement standard deviation, σ_k . The adjusted interlaboratory sample standard deviation is the sample standard deviation, s_k , multiplied by the bias-correction factor, a'_n , found in Table 1. In this Practice, all references to computed and fitted values of the interlaboratory sample standard deviation refer to adjusted values.

(c) Plot s_k versus T_k .

(d) Using ordinary least squares (OLS, see Caulett and Boddy (5)), regress s_k on T_k , temporarily assuming that the Straight-line Model is valid. The regression provides coefficients, g and h , in the relationship,

$$s_k = g + hT_k + \text{error}. \quad (4)$$

Compute residuals,

$$r_k = s_k - (g + hT_k) \quad (5)$$

Plot r_k versus T_k .

(e) Evaluate the reasonableness of the Constant Model for the ILSD (Model A) as follows: First, note the p -value associated with slope estimate, h , from the OLS regression. If the p -value is less than 5 %, there is statistically significant slope, and the Constant model should be rejected; proceed to the next step. Second, examine the plots produced in (c) and (d). If obvious systematic curvature is present (for example, quadratic-like behavior), both the Constant Model and the Straight-line Model should be rejected; proceed to (i). If the Constant Model is not rejected, proceed to 6.3.4.

(f) The Constant Model (Model A), has been rejected because of statistically significant slope. Evaluate the reasonableness of the Straight-line model for the ILSD (Model B). Examine the plot produced in (d). If obvious systematic curvature is present (for example, quadratic-like behavior), with a minimum that appears to be in the concentration range, the Straight-line Model should be rejected; proceed to (j). If the Straight-line Model is not rejected by this examination, proceed to 6.3.4, or, optionally, conduct a formal test for curvature, as follows in (g) through (i) (note that the usual and more general lack-of-fit test is not applicable for this modeling effort because there are no replicate sample standard deviations, s_k , for any concentration).

(g) Using OLS, regress T_k^2 on T_k , producing fitted coefficients u and v , used only to compute residuals, q_k , which comprise the orthogonal component of the quadratic term, T_k^2 :

$$q_k = (\text{predicted } T_k^2) - T_k^2 = (u + vT_k) - T_k^2 \quad (6)$$

(h) Using OLS, regress s_k on T_k and q_k simultaneously, producing fitted coefficients g and h (as before), but additionally Q :

$$s_k = g + hT_k + Qq_k + \text{error} \quad (7)$$

The only results of interest are the statistical significance and the sign of Q . These results collectively indicate the strength of evidence for curvature.

(i) Note the p -value, p_Q , associated with Q . Because q_k is orthogonal to T_k , this p -value indicates the level of statistical significance of (quadratic) curvature.

NOTE 3—Even though the test for curvature uses a quadratic term, a quadratic model is not one of the three recommended model choices. If $p_Q < 5\%$ and $Q > 0$, there is sufficient statistical evidence of curvature in the relationship between s_k and T_k to warrant the use of the Hybrid Model, Model C ($Q > 0$ ensures that the increase in s_k with respect to T_k is faster than linear). If these conditions do not hold, then the Straight-line Model (Model B) is the appropriate model to use. Proceed to 6.3.4

(j) The Hybrid Model for the ILSD (Model C) can be used if there is evidence of curvature.

(k) To evaluate the reasonableness of the Hybrid Model, Model C, the model must first be fitted using nonlinear least squares (NLLS), either by Newton's-Method iteration (presented in the appendix), or another NLLS method.

(l) The fit from the Hybrid Model should be evaluated. A plot of the residuals, in log form, should be constructed: plot r_k versus T_k , where:

TABLE 1 Bias-Correction Adjustment Factors for Sample Standard Deviations Based on n Measurements (at a particular concentration)^A

n	2	3	4	5	6	7	8	9	10
a'_n	1.253	1.128	1.085	1.064	1.051	1.042	1.036	1.031	1.028

^AFor each true concentration, T_k , the adjusted value $s_k = a'_n s'_k$ should be modeled in place of sample standard deviation, s'_k . For $n > 10$, use the formula, $a'_n = 1 + [4(n-1)]^{-1}$. See Johnson and Kotz (7).

$$r_k = \ln s_k - \ln \hat{s}_k \quad (8)$$

and \hat{s}_k is the predicted value of s_k using the model. The plot should show no systematic behavior (for example, curvature). If the fit satisfies both types of evaluation, go to 6.3.4. Otherwise, a different (and possibly more complex) model may be used, such as the exponential model: $s = g \exp \{hT\} \cdot (1 + \text{error})$. If there are enough true concentrations, a model with more coefficients could be considered; possibilities include quadratic (strictly increasing with increasing concentration), or even cubic.

6.3.4 *Fit the Mean-Recovery Model*—The mean-recovery model is a simple straight line,

$$\text{Model R: } Y = a + bT + \text{error} \quad (9)$$

The fitting procedure depends on the model selection from 6.3.3. If the constant model, Model A, was selected for ILSD, then OLS can be used to fit Model R for mean recovery (see the left column of Table 2, or Caulcutt and Boddy (5)). If a nonconstant ILSD model was selected, such as the Straight-line Model (Model B), or the Hybrid Model (Model C), then weighted least squares (WLS) should be used to fit mean recovery. The WLS approximately provides the minimum-variance unbiased linear estimate of the coefficients, a and b . The WLS procedure is described in 6.3.4.1

6.3.4.1 *Weighted Least Squares Procedure, Using the Interlaboratory Standard Deviation (ILSD) Model:*

(a) Using the ILSD model and coefficient estimates from 6.3.3, compute the predicted interlaboratory standard deviation, \hat{s}_k , for each true concentration, T_k :

$$\text{Model B: } \hat{s}_k = g + hT_k \quad (10)$$

$$\text{Model C: } \hat{s}_k = (g^2 + [hT_k]^2)^{(1/2)} \quad (11)$$

(b) Compute weights for WLS:

$$w_k = (\hat{s}_k)^{-2} \quad (12)$$

Note that if WLS is carried out using computer software, the default setting for weights may be different. For example, instead of supplying the values, $(\hat{s}_k)^{-2}$, as weights, the software may require the user to supply values (\hat{s}_k) or $(\hat{s}_k)^2$ as weights that are internally transformed by the software.

TABLE 2 Ordinary Least Squares (OLS) and Weighted Least Squares (WLS) Computations to Estimate Straight-line Model Coefficients

(Computations shown for convenience and contrast)	OLS	WLS
	$\bar{T} = \frac{1}{n} \sum_{i=1}^n T_i$	$\bar{T}_w = \frac{1}{n} \sum_{i=1}^n w_i T_i$
	$\bar{y} = \frac{1}{n} \sum_{i=1}^n y_i$	$\bar{y}_w = \frac{1}{n} \sum_{i=1}^n w_i y_i$
	$S_{TT} = \sum_{i=1}^n (T_i - \bar{T})^2$	$S_{wTT} = \sum_{i=1}^n w_i (T_i - \bar{T})^2$
	$S_{TY} = \sum_{i=1}^n (T_i - \bar{T})(y_i - \bar{y})$	$S_{wTY} = \sum_{i=1}^n w_i (T_i - \bar{T})(y_i - \bar{y})$
	slope = $b = S_{TY} / S_{TT}$	slope = $b = S_{wTY} / S_{wTT}$
	intercept = $a = \bar{y} - b\bar{T}$	intercept = $a = \bar{y}_w - b\bar{T}_w$

(c) Carry out WLS computations analogous to OLS computations. See Table 2 or Caulcutt and Boddy (5). The result will be coefficient estimates, a and b , for the mean-recovery model, Model R. Appendix II describes three approximate approaches to WLS commonly practiced, but not acceptable for this application.

(d) After fitting, the mean-recovery model should be evaluated for reasonableness and lack of fit. This evaluation should be done by ensuring the following: (1) The fit is statistically significant (overall p -value $< 5\%$); (2) The lack-of-fit p -value (if available; see Caulcutt and Boddy (5) or Draper and Smith (4)) is not statistically significant (lack-of-fit p -value $> 5\%$); (3) A plot of the residuals shows no obvious systematic curvature (for example, quadratic-like behavior). If the mean-recovery model fails the evaluation, then the study supervisor will have to determine if only a subset of the data should be analyzed (perhaps the model fails for the higher concentration(s)), or if more data are needed.

6.4 *Compute the IQE*—The IQE is computed using the ILSD model to estimate the interlaboratory standard deviation, and using the mean-recovery model to scale the standard deviation. For any computed IQE to be valid, it must lie within the range of concentrations used in the study. The general form of the computation is to find the solution, LQ (within the range of concentrations used in the study), to the following equation:

$$T = (100/Z) \cdot G(T) \quad (13)$$

where function $G(T)$ is the estimated interlaboratory standard deviation (in concentration units) of true value, T , and Z is taken to be 10, 20, or 30, in increasing order. That is, the first attempt is to compute $\text{IQE}_{10\%}$. If $\text{IQE}_{10\%}$ does not exist or is outside the range of concentrations used in the study, then $\text{IQE}_{20\%}$ is computed, if possible. If $\text{IQE}_{20\%}$ does not exist or is outside the range of concentrations used in the study, then $\text{IQE}_{30\%}$ is computed, if possible. If appropriate for a particular use, $\text{IQE}_Z\%$ can be computed for any value of $Z < 10$, but $Z > 30$ is not recommended. Thus, the IQE computation depends on the form of the ILSD model, which is part of function G . The ratio, $Z' = 100 \cdot h/b$, represents the limit of the %RSD achievable. Therefore the strictest IQE achievable by the analytical method studied is $\text{IQE}_{Z'}\%$. For example, if $Z' = 100 \cdot 0.17/1.0 = 17$, then the strictest IQE achievable would be the $\text{IQE}_{20\%}$ (according to the nearest higher multiple of 10).

6.4.1 *ILSD Constant Model (Model A)*—In this case, $\hat{s} = g$; hence $G(T) = g/b$ and $\text{LQ} = (100/Z) \cdot g/b$. Thus,

$$\text{IQE}_{Z\%} = (100/Z) \cdot g/b \quad (14)$$

6.4.2 *ILSD “Straight-line” Model (Model B)*—In this case, $\hat{s} = g + hT$; hence $G(T) = (g + hT)/b$. To find the IQE, one must solve for T : $T = (100/Z) \cdot (g + hT)/b$. The solution is:

$$\text{IQE}_{Z\%} = g / (b \cdot (Z/100) - h) \quad (15)$$

6.4.3 *ILSD Hybrid Model (Model C)*—(additive and multiplicative error, in accordance with Locke and Lorenzato (3)). In this case, $\hat{s} = (g^2 + [h \cdot T]^2)^{(1/2)}$; hence $G(T) = (g^2 + [h \cdot T]^2)^{(1/2)}/b$. To find the IQE, one must solve

$$T = ((100/Z)/b) (g^2 + [h \cdot T]^2)^{(1/2)} \quad (16)$$

This solution is derived by squaring each side of the equation and solving to get: $IQE_{Z\%} = g / [(b \cdot Z/100)^2 - h^2]^{1/2}$, where the positive square root is taken.

6.5 *Non-trivial Amount of Censored Data*—More than 10 % of the data for at least one true concentration may have been reported as nondetects or less-thans. Despite the attempt in 6.2.3.1 to reduce or eliminate reported nondetects or less-thans, they may still occur at a level that disrupts the data analysis presented in 6.3 and 6.4. If there is excessive censoring, the study supervisor should contact laboratories with such measurements to see whether the data can be extracted in uncensored form from data archives. If this effort is not a sufficient remedy, serious consideration should be given to augmenting the IQE study with measurements of samples at new and different concentrations (generally, higher).

7. Data Analysis

7.1 The data analysis for eliminating data is given in Section 10 of Practice D 2777.

7.2 The data analysis involved in computing an IQE is shown by example in Section 10 of this practice.

8. Report

8.1 The analysis report should be structured as in Annex A1.

8.1.1 The report should be given a second-party review to verify that:

8.1.1.1 The data transcription and reporting have been performed correctly,

8.1.1.2 The analysis of the data has been performed correctly, and

8.1.1.3 The results of the analysis have been used appropriately, including assessment of assumptions necessary to compute an IQE.

8.1.2 A statement of the review and the results should accompany the report. Reviewer(s) should be qualified in one or both of the following areas: (1) applied statistics, and (2) analytical chemistry.

9. Rationale

9.1 The basic rationale for the IQE is contained in Currie (1). The IQE is a performance characteristic of an analytical method, to paraphrase Currie. As with the Interlaboratory Detection Estimate (IDE) (described in Practice D 6091), the IQE is vital for the planning and use of chemical analyses. The IQE is another benchmark indicating whether the method can adequately meet measurement needs.

9.2 The idealized definition of $IQE_{Z\%}$ is that it is the lowest concentration, LQ, that satisfies: $T = (100/Z) \sigma_T$ (where σ_T is the actual standard deviation of interlaboratory measurements at concentration T), which is equivalent to satisfying, $\%RSD = \sigma_T/T = Z\%$. In other words, $IQE_{Z\%}$ is the lowest concentration with Z % RSD (assuming such a concentration exists). If, as is commonly the case, %RSD declines with increasing true concentration, then the relative uncertainty of any measurement of a true concentration greater than the IQE will not exceed $\pm Z\%$. The range, $\pm 3\sigma_{LQ}$, is an approximate prediction or confidence interval very likely to contain the measurement, which is assumed to be Normally distributed. This assertion is

based on critical values from the Normal distribution (or from the Student's *t* distribution if σ is estimated rather than known). Then, with high confidence, the relative error of any measurement of a true concentration greater than the IQE will not exceed $\pm 3Z\%$. For example, a measurement above the $IQE_{10\%}$ (and assumed to have true concentration above the IQE) could be reported as 6 ppb ($\pm 30\%$) = 6 (± 2) ppb, with a high degree of certainty.

9.3 There are several real-world complications to this idealized situation. See Maddalone et al (8), Gibbons (9), and Coleman et al (10). Some of these complications are listed as follows:

9.3.1 Analyte recovery is not perfect; the relationship between measured values of concentrations and true concentrations cannot be assumed to be trivial. There is bias between true and measured values. Recovery can and should be modeled. Usually a straight line will suffice.

9.3.2 Variation is introduced by different laboratories, analysts, models and pieces of equipment; environmental factors; flexibility/ambiguity in a test method; contamination; carry-over; matrix influence; and other factors. It is intractable to model these factors individually, but their collective contributions to measurement ILSD can be observed, if these contributions are part of how a study is designed and conducted.

9.3.3 The interlaboratory standard deviation of measurements is generally unknown, and may change with true concentration, possibly because of the physical principle of the test method. To ensure that a particular %RSD is attained at or above the IQE, there must be a way to predict the ILSD at different true concentrations. Short of severely restricting the range of concentrations for a study, prediction is accomplished by an empirical ILSD model. In all of the respects discussed in 9.1-9.3, $IQE_{10\%}$ is similar to the AML developed by Gibbons et al (11). However, the AML follows an approximate approach, where the standard deviation used in the 10σ formula is estimated at a detection critical value, and then is taken to be a constant (over a trace-level range of concentrations) for the 10σ computation. In contrast, $IQE_{10\%}$ follows the "more statistically and conceptually rigorous" approach described by Gibbons et al (9), and contained in Currie (1). This greater rigor comes at the risk of: (a) possibly being unattainable for some methods (for which only a less strict level of %RSD can be ensured); (b) having uncertainty that is potentially complex, and depends both on the model used and on the data.

10. Example (Hybrid ILSD Model)

10.1 *Identify and Fit the ILSD Model*—Ten laboratories participated in a (synthesized) IQE study, where single measurements were made at each of seven concentrations: $T_k = \{0.0, 0.50, 1, 2, 4, 8, 12\}$ ppb. Considerations of 6.2, 6.3.1, and 6.3.2 are not described here. The procedure described in 6.3.3 is followed, assuming that no data were eliminated in accordance with 6.3.2.

10.1.1 The reported measurements are shown in Table 3. These values are also shown in Fig. 2. The straight-line recovery model appears to be plausible, and the data appear to have measurement ILSDs that increase with concentration.

TABLE 3 Reported Measurements and Computed Statistics from the Example IQE Study

True Concentration T_k , ppb	Reported Measurement Values, y_k , One per Laboratory for each T_k	$s_k =$ Interlaboratory Sample Standard Deviation (adjusted)	$\ln s_k$	$\hat{s}_k =$ Predicted Standard Deviation (Hybrid Model)	WLS Weights: $w_k = (\hat{s}_k)^{-2}$	q_k , Orthogonal Component of $(T_k)^2$
0	-0.105, 0.263, 0.293, 0.187, 0.106, 0.329, 0.080, 0.524, 0.278, 0.206	0.1729	-1.7549	0.1840	20.54	13.029
0.5	0.354, 0.724, 0.682, 0.327, 0.527, 0.868, 0.730, 0.434, 0.794, 0.642	0.1929	-1.6454	0.1927	26.93	7.453
1	1.241, 0.668, 1.200, 1.370, 1.106, 0.964, 0.949, 1.421, 1.032, 1.134	0.2270	-1.4829	0.2168	21.28	2.376
2	2.174, 2.388, 2.153, 2.366, 2.306, 2.309, 1.663, 2.841, 1.933, 1.809	0.3449	-1.0644	0.2939	11.58	-6.277
4	3.660, 3.734, 3.167, 3.578, 4.278, 3.383, 3.873, 4.479, 3.919, 3.856	0.3995	-0.9175	0.4940	4.10	-17.582
8	6.592, 7.520, 6.822, 7.751, 7.771, 7.296, 8.578, 6.863, 7.840, 8.821	0.7521	-0.2849	0.9351	1.14	-16.194
12	9.496, 9.081, 13.942, 10.547, 9.324, 13.148, 10.994, 11.774, 12.320, 13.521	1.8519	0.6162	1.3875	0.52	17.194

10.1.2 Interlaboratory sample standard deviations at each true concentration are computed, adjusted to remove bias (using Table 1) and are shown in Table 3.

10.1.3 A plot of interlaboratory sample standard deviation versus true concentration is shown in Fig. 3. The plot provides additional qualitative evidence of an increase in standard deviation with increasing concentration.

10.1.4 A straight-line regression (using OLS) is conducted of the interlaboratory sample standard deviations, s_k , versus T_k . The results are shown in Table 4, and the fit is shown in Fig. 3.

10.1.5 The slope estimate, h , is statistically significant with a p -value of $0.0012 < 5\%$, so the Constant ILSD model (Model A) is rejected.

10.1.6 The reasonableness of the straight-line model (Model B) is evaluated using the lower plot (in Fig. 3) the plot of residuals versus true concentration. There is subjective appearance of systematic curvature (a roughly U-shape to the residuals).

10.1.7 To assess more formally the need for a model with curvature (Hybrid, Model C) instead of the straight line model (Model B), a formal test is conducted.

10.1.7.1 Using OLS, $(T_k)^2$ is regressed on T_k , producing residuals, q_k , shown in Table 3.

10.1.7.2 Using OLS, s_k is regressed on T_k and q_k together, once again producing estimates of coefficients g and h , and additionally Q , the coefficient of q . The results are shown in Table 5.

10.1.7.3 From Table 5, it can be seen that $p_Q = 0.0096 < 5\%$, and $Q = 0.013 > 0$, so there is sufficient evidence of curvature to warrant using the Hybrid Model (Model C).

10.1.8 Model C, the Hybrid Model, is used to fit the sample standard deviation data in Table 3, using NLLS solved by Newton's-method iteration, as presented in the appendix. The steps are as follows:

10.1.8.1 Compute the natural log sample standard deviation, $\ln s_k$, for each true concentration, T_k . See Table 3.

10.1.8.2 Let j be the index of iteration, and set $j=0$. Compute initial values, g_0 and h_0 , as follows:

$$g_0 = s_1 = 0.173 \quad (17)$$

$$h_0 = (s_{max} - s_1)/(T_{max} - T_1) = 0.140$$

See Table 6.

10.1.8.3 Compute the natural log of the estimated standard deviation, $\ln s_{sk}$, for each k , using the current values of g_j and h_j (not shown).

10.1.8.4 Compute the residuals $r_k = \ln s_k - \ln s_{sk}$ for each k (not shown).

10.1.8.5 Compute $fg_k = g_j / \exp\{2 \ln s_{sk}\}$ for each k (not shown).

10.1.8.6 Compute $fh_k = h_j (T_k)^2 / \exp\{2 \ln s_{sk}\}$ for each k (not shown).

10.1.8.7 Compute intermediate statistics: u , v , c , d , p , and q . See Table 6.

10.1.8.8 Compute the j th-step changes to g and h (see Table 6):

$$\begin{aligned} \Delta g &= d(Vp - cq) & dg\% &= 100 |\Delta g / g| \\ \Delta h &= d(Uq - cp) & dhT\% &= 100 |\Delta h| T_{max} \end{aligned}$$

10.1.8.9 Compute the new g and h (see Table 6):

$$g_{j+1} = g_j + \Delta g \quad h_{j+1} = h_j + \Delta h$$

10.1.8.10 Iterate (increase j by 1, and return to 10.1.8.3) until $dg\% < 1\%$ and $dhT\% < 1\%$.

10.1.8.11 As can be seen in Table 6 for $j=2$, $dg\%=0.02\% < 1\%$ and $dhT\%=0.2\% < 1\%$, so convergence is achieved after the second step of iteration, with $g=0.184$ and $h=0.1146$.

As seen in Table 3, the coefficients, g and h , are used to compute a predicted measurement standard deviation, \hat{s}_k , at each of the T_k values. The \hat{s}_k values are then used to compute weights, w_k , also seen in Table 3.

10.1.9 WLS uses the weights, w_k , to fit the straight-line mean-recovery function. The results are shown in Table 7.

10.1.10 Finally, $IQE_{Z\%}$ is computed. First, the lowest achievable %RSD is estimated, in accordance with the formula for the Hybrid Model: $Z' = 100h/b = 100 \cdot 0.1146 / 0.931 = 12$, which is rounded up to a whole-number multiple of 10: $Z=20$. Hence, $IQE_{20\%}$ can be computed (but not $IQE_{10\%}$), as follows: $IQE_{20\%} = g / [(b \cdot 20 / 100)^2 - h^2]^{(1/2)} = 0.184 / [(0.931 \cdot 20 / 100)^2 - 0.1146^{(1/2)}] = 1.254$ ppb

For comparison purposes, a simple, model-free quantitation limit equal to five times the sample measurement standard deviation from blank replicates might be $5 \cdot s_1 = 5 \cdot 0.173 = 0.865$




FIG. 2 Reported Concentration Measurement (ppb) Versus True Concentration (ppb), One Measurement per Laboratory at Each Concentration With Weighted Least Squares Straight-Line Fit and (Below) Residuals

ppb. This estimate would be even lower if an intralaboratory standard deviation were used instead of an interlaboratory standard deviation.

10.1.11 It is also possible to compute $IQE_{30\%}$, as follows:

$$IQE_{30\%} = g / [(b \cdot 30 / 100)^2 - h^2]^{(1/2)} = 0.184 / [0.931 \cdot 30 / 100)^2 - 0.1146^2]^{(1/2)} = 0.722 \text{ ppb}$$

10.1.12 The $IQE_{20\%}$ value can be confirmed by means of a graphical device: plotting both sides of Eq 13, $(T = (100/Z) \cdot G(T))$, versus true concentration, (T) , as seen in Fig. 4. These functions will intersect at $T = IQE_{Z\%}$. Recall from 6.4 that $10 G(LQ) = (100/Z) \cdot (\text{estimated interlaboratory standard deviation}) = (100 / Z) \cdot \hat{s}_k / b$. The scaling by b transforms the estimates to true concentration units (in contrast to measured units). The two curves intersect at the true concentration that

satisfies Eq 13, seen to be slightly more than 1 ppb. Also shown are the suitably scaled sample standard deviations, $(100/Z) \cdot \hat{s}_k / b$, so the goodness of fit for the Hybrid Model can be seen.

10.1.13 Fig. 4 suggests a way to approximate graphically $IQE_{20\%}$, without explicit ILSD modeling and with limited computation. Graphically estimating the IQE from Fig. 3 involves locating the point of intersection between two rising curves, a process that would be made more difficult without the fitted ILSD model. If, instead, the difference, $D = T - (100/Z) \cdot G(T)$ (taken from the two sides of Eq 13) is plotted versus T , and compared to zero, graphical resolution is enhanced,




FIG. 3 Sample Standard Deviations (+) Versus True Concentration, with Straight Line Fit, Hybrid Model Fit, and Residuals from Straight Line Fit (Lower Plot), All in ppb

TABLE 4 Straight-Line OLS Fit of s on T

$$\text{Standard deviation} = s = g + hT = 0.06498 + 0.12678 T$$

Summary of Fit	
RSquare	0.896432
RSquare Adj	0.875719
Root Mean Square Error	0.212178

Term	Parameter Estimates			
	Estimate	Standard Error	t Ratio	Prob> t
g (Intercept)	0.0649765	0.110288	0.59	0.5814
h (slope)	0.1267813	0.019272	6.58	0.0012

TABLE 5 Summary of OLS Fit of s on T and q

Term	Estimate	SI	t Ratio	Prob > t
g (Intercept)	0.0649765	0.048621	1.34	0.2524
h (slope w.r.t. T)	0.1267813	0.008496	14.92	0.0001
Q (coefficient of q)	0.0129282	0.002774	4.66	0.0096

especially if sample standard deviations are used rather than standard deviations estimated from the ILSD model. See Fig. 5.

10.1.13.1 To make an unbiased estimate of the IQE, one must scale by an estimate, b' , of the mean-recovery slope,

estimated more precisely by b in the IQE procedure. One can use OLS to fit y on T , or (less preferable) compute $b' = [\text{mean}(y_{T_{\max}}) - \text{mean}(y_0)] / [T_{\max} - T_1]$, where $\text{mean}(y_{T_{\max}})$ is the mean of all y measurements made at the highest concentration, T_{\max} . For the data in this example, Fig. 5 shows the plot of the difference, $D = T - (100/Z) \cdot s/b'$ versus T , where b' was obtained by OLS. Without an ILSD model, it is not obvious where $D = 0$, but it appears to be at approximately $T = 1.25$ ppb. In general, this approximate approach cannot always be




FIG. 4 Plot of True Standard Deviation Concentration, T (Straight Line), $5 \cdot$ (Predicted Standard Deviation), and $5 \cdot$ (Sample Standard Deviation) Versus T - Providing Graphical Confirmation of $IQE_{20\%}$ Result




FIG. 5 Plot of $[D = \text{True Concentration } (T) \text{ Minus } 5 \cdot (\text{Sample Standard Deviation})]$ Versus T - Providing Graphical Estimation of $IQE_{20\%}$ Where $D = 0$ (For reference, $5 \cdot (T - \text{fit})$ curve is also shown). All in ppb.

TABLE 6 Summary Statistics from Newton's Method Fit of Hybrid Model

j	g	h	u	v	c	d	p	q	Δg	Δh	$dg\%$	$dh T\%$
0	0.173	0.1400	73.11	176.87	27.77	8.22E-05	0.0634	-4.3765	0.0109	-0.0265	6.3	227
1	0.183	0.1135	74.99	238.43	33.28	5.96E-05	0.0540	0.2681	0.0002	0.0011	0.1	11.5
2	0.184	0.1146	74.47	234.83	32.89	6.10E-05	-0.0016	0.0037	-3E05	2E-05	0.02	0.2

relied on to produce an estimate of $IQE_{Z\%}$ because sample standard deviations are noisy, and the plot of D versus T may not even be monotonic near $D = 0$.

11. Keywords

11.1 critical limit; matrix effects; precision; quantitation; quantitation limit



TABLE 7 WLS Straight-Line Fit for Measured Values Versus True

Measurement = $y = 0.19399 + 0.93062T$
 Measurement = $y = a + b T$

Summary of Fit				
	RSquare			0.963246
	RSquare Adj			0.962706
	Root Mean Square Error			0.994013
Parameter Estimates				
Term	Estimate	Standard Error	t Ratio	Prob> t
a (intercept)	0.1939874	0.038359	5.06	< 0.0001
b (slope)	0.9306236	0.022045	42.22	< 0.0001

ANNEX

(Mandatory Information)

A1. ANNOTATED OUTLINE FOR ANALYSIS REPORTS

A1.1 This outline presents the information to be included in the reports of analysis performed in accordance with this practice.

A1.2 Single-Laboratory IQE Report:

A1.2.1 Identification of laboratory, analytical method, analyte(s), matrix (or matrices), and sample properties (for example, volume).

A1.3 Any anomalies in the study, including QA/QC sample results.

A1.4 Interlaboratory Quantitation Estimate Report:

A1.4.1 Data-screening results, individual values and laboratories omitted from further analysis, and missing values.

A1.4.2 ILSD model selected.

A1.4.3 Coefficient estimates for the ILSD model and mean-recovery model.

APPENDIXES

(Nonmandatory Information)

X1. FITTING THE HYBRID (ROCKE AND LORENZATO (3)) MODEL FOR ANALYTICAL MEASUREMENTS, USING NEWTON'S METHOD OF NON-LINEAR LEAST SQUARES (NLLS)

X1.1 The following numerical procedure can be conveniently carried out by using computer spreadsheet software:

X1.1.1 Initialize: The index, j , is the step number for iteration. Set $j=0$.

X1.1.1.1 Compute the natural log of the sample standard deviation, ls_k , for each true concentration, T_k .

NOTE X1.1—The log transformation standardizes the residuals so that the sum of squares of logs of relative errors is minimized. Log-relative errors are preferred to absolute errors, since the latter are almost certainly unequal in variation.

X1.1.1.2 Compute initial values, g_0 and h_0 , as follows:

$g_0 = s_1$ (the sample standard deviation for the lowest concentration, T_1 ; usually $T_1=0$)
 $h_0 = (s_{max} - s_1) / (T_{max} - T_1)$ if $s_{max} > s_1$, where s_{max} is the maximum sample standard deviation of measurements, made at concentration, T_{max} ; Otherwise, set $h_0 = 0$.

X1.1.1.3 Compute the natural log of the estimated standard deviation, lss_k , for each T_k , using the current estimates, g_j and h_j :

$$lss_k = f(T_k), \text{ where we define } f(T_k) = \ln \sqrt{g_j^2 + h_j^2 T_k^2} \quad (X1.1)$$

X1.1.1.4 Compute the difference (residual), r_k , between the log sample standard deviation and estimated log standard deviation for each k :

$$r_k = ls_k - lss_k \quad (X1.2)$$

Note that r_k is the natural log of the ratio of the sample standard deviation to the estimated standard deviation, so r_k represents log-proportional error, and is ideally equal to zero.

X1.1.1.5 Compute fg_k , the slope (that is, numerical derivative) of $f(T)$ with respect to g , for each k :

$$fg_k = g_j / \exp\{2 lss_k\} \quad (X1.3)$$

X1.1.1.6 Compute fh_k , the slope of $f(T)$ with respect to h , for each T_k :

$$fh_k = h_j (T_k)^2 / \exp\{2 lss_k\} \quad (X1.4)$$

X1.1.1.7 Compute the following intermediate statistics:

$$u = \sum_k (fg_k)^2 \quad v = \sum_k (fh_k)^2 \quad c = \sum_k (fg_k \cdot fh_k)$$



$$d = \frac{1}{uv - c^2} \quad p = \sum_k (fg_k \cdot r_k) \quad q = \sum_k (fh_k \cdot r_k) \quad g_{j+1} = g_j + \Delta g \quad (X1.5)$$

$$h_{j+1} = h_j + \Delta h$$

X1.1.1.8 Compute the j th step changes to g and h (made to reduce the sum of squared residuals), and % relative changes:

$$\Delta g = d (vp - cq) \quad dg\% = 100 |\Delta g / g_j|$$

$$\Delta h = d (uq - cp) \quad dhT\% = 100 |\Delta h / h_j|_{T_{max}}$$

X1.1.1.9 Compute new g and h estimates:

X1.1.1.10 If $dg\% < 1\%$ and $dhT\% < 1\%$, then stop and use g_{jh} and h_{jh} as the final estimates. Otherwise, increase j by 1, and go to X1.1.1.3.

X2. THREE WLS APPROXIMATIONS TO BE AVOIDED

X2.1 There are three approximate approaches to WLS commonly used, but not acceptable for this practice. One approach uses the reciprocal-squared sample standard deviations, s_k^{-2} , as weights. Since this practice involves the explicit evaluation and selection of a standard-deviation model, the predicted value for s_k is probably more precise than a sample value, and the former value should be used to compute weights. A second approach omits the blank measurements, and divides the rest of the measurements by the true concentrations. Then, OLS is carried out, using the independent variable, $1/T$, in the following model:

$$Y/T = a (1/T) + b + \text{error}. \quad (X2.1)$$

This approach is not acceptable because it leads to loss of data and because the weights so generated implicitly assume that interlaboratory standard deviation is strictly proportional to true concentration. A proportional relationship cannot hold for arbitrarily small concentrations. The third approach exploits the same approximate (but untrue) proportional relationship to obtain mathematically simpler WLS formulas.

X3. GLOSSARY OF KEY SYMBOLS, ACRONYMS, AND LABELS

σ —true interlaboratory standard deviation
 Δg —one iteration's change in the estimate of g , the intercept coefficient in the Hybrid Model
 Δh —one iteration's change in the estimate of h , the slope coefficient in the Hybrid Model
 a —estimate of the slope in the mean-recovery curve (Straight-line Model)
 a'_{r} —adjustment factor used to remove bias from the sample interlaboratory standard deviation
 AML—Alternative Minimum Level, a quantitation limit that is similar to the IQE (and compatible in approach)
 b —estimate of the slope in the mean-recovery curve (Straight-line Model)
 b' —crude estimate of b
 c —intermediate variable used in estimating g and h for the Hybrid model, by nonlinear least squares. Similar to d , p , q , u , and v
 D —difference between T and $((100/Z) \cdot (\text{estimated interlaboratory standard deviation}))$, used for approximate, graphical determination IQE
 d —similar to c
 $f(T)$ —the natural log of the current estimate of the interlaboratory standard deviation at concentration, T
 g —estimate of the intercept in the Hybrid model of interlaboratory standard deviation
 $G(T)$ —the (generic) model of the interlaboratory standard deviation
 g_o —initial estimate of g
 IDE—the interlaboratory detection estimate, defined and described in Practice D 6091
 ILSD—interlaboratory standard deviation
 $\text{IQE}_{Z\%}$ —interlaboratory quantitation estimate associated with approximately $Z\%$ RSD
 j —iteration index used for nonlinear least squares solution of the coefficients for the Hybrid Model for ILSD
 k —index used for different concentrations, T_k , and associated statistics
 LQ—Another designation for the IQE, in accordance with Currie's notation
 Model A—Constant model for ILSD
 Model B—Straight-line model for ILSD; interlaboratory standard deviation increases with increasing concentration

Model C—Hybrid model for ILSD; combines additive and multiplicative error, with interlaboratory standard deviation that increases with increasing concentration, according to the model proposed by Rocke and Lorenzato
 Model R—the straight-line model for the mean-recovery curve
 NLLS—nonlinear least squares, where coefficients in a nonlinear model are computed to minimize the sum of the squares of the residuals (that is, the differences between the predicted and actual values)
 OLS—ordinary least squares, a fitting technique for a linear (that is, additive) model that minimizes the sum of the squares of the residuals (that is, the differences between predicted and actual values)
 p —similar to c
 q —similar to c
 q_k — k^{th} value in the T^2 quadratic component that is orthogonal to T
 Q —intermediate variable used in ILSD model selection, to test for statistically significant curvature
 QL—quantitation limit (also called practical quantitation limit, PQL); see LQ
 r —the estimated lowest limit of %RSD achievable, based on study results, for a particular measurement system, matrix, and analyte
 r_k (unrelated to r)—the residual associated with T_k from a precision-model fit; defined as the difference in log sample standard deviation and log estimated (predicted) standard deviation
 RSD—relative standard deviation, that is, the standard deviation divided by the concentration, (both generally estimated)
 s —modeled value of the interlaboratory standard deviation, including error
 s_k —sample interlaboratory standard deviation at true concentration, T_k , adjusted to remove bias
 s_{max} —maximum sample ILSD: equal to $\max \{s_1, s_2, \dots\}$
 T —true concentration
 T_k — k^{th} value of true concentration in the study
 T_{max} —maximum concentration in the study; equal to $\max \{T_1, T_2, \dots\}$
 WLS—weighted least squares, a modified form of ordinary least squares. WLS incorporates nonuniform variability in the data
 Y —random variable representing a reported measurement
 Z —level of RSD



D 6512 – 00

REFERENCES

- (1) Currie, L., "Nomenclature in Evaluation of Analytical Methods Including Detection and Quantification Capabilities," *Pure and Applied Chemistry*, Vol 67, 1995, pp. 1699–1723.
- (2) Oppenheimer, L., Capizzi, T.P., Weppelman, R.M., and Mehta, H., "Determining the Lowest Limit of Reliable Assay Measurements," *Analytical Chemistry*, Vol 55, 1983, pp. 638–643.
- (3) Rocke, D.M., and Lorenzato, S., "A Two-Component Model for Measurement Error in Analytical Chemistry," *Technometrics*, Vol 37, No. 2, May 1995, pp. 176–184.
- (4) Draper, N., and Smith, J., *Applied Regression Analysis*, 2nd ed., John Wiley & Sons, Inc., New York, NY, 1981, pp. 33–40.
- (5) Caulcutt, R., and Boddy, R., *Statistics for Analytical Chemists*, Chapman and Hall, New York, NY, 1983, pp. 79–85, 100–110, 114–123.
- (6) Carroll, R., and Ruppert, D., *Transformation and Weighting in Regression*, Chapman and Hall, New York, NY, 1988, pp. 10–11, 62–69.
- (7) Johnson, N., and Kotz, S., *Continuous Univariate Distributions*, Vol 1, John Wiley & Sons, Inc., New York, NY, 1970, p. 63.
- (8) Maddalone, R., Rice, J., Edmondson, B., Nott, B., and Scott, J., "Defining Detection and Quantitation Levels," *Water Environment and Technology*, Vol 5, No. 3, January 1993, pp. 41–44.
- (9) Gibbons, R., *Statistical Methods for Groundwater Monitoring*, John Wiley & Sons, Inc., New York, NY, 1994, pp. 95–121.
- (10) Coleman, D., Auses, J., and Grams, N., "Regulation - From an Industry Perspective or Relationships Between Detection Limits, Quantitation Limits, and Significant Digits," *Chemometrics and Intelligent Laboratory Systems*, Vol 37, 1997, pp. 71–80.
- (11) Gibbons, R., Coleman, D., and Maddalone, R., "An Alternative Minimum Level Definition for Analytical Quantification," *Environmental Science and Technology*, Vol 31, 1997, pp. 2071–2077.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).