



Designation: D 1209 – 9700

Standard Test Method for Color of Clear Liquids (Platinum-Cobalt Scale)¹

This standard is issued under the fixed designation D 1209; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope

1.1 This test method describes a procedure for the visual measurement of the color of essentially light colored liquids (Note 1). It is applicable only to materials in which the color-producing bodies present have light absorption characteristics nearly identical with those of the platinum-cobalt color standards used.

NOTE 1—A procedure for estimating color of darker liquids, described for soluble nitrocellulose base solutions, is given in Methods D 365.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements see Section 6.

1.3 For specific hazard information, see the Material Safety Data Sheet.

2. Referenced Documents

2.1 ASTM Standards:

D 156 Test Method for Saybolt Color of Petroleum Products (Saybolt Chromometer Method)²

¹ This test method is under the jurisdiction of ASTM Committee D-1 on Paint and Related Coatings, Materials, and Applications and is the direct responsibility of Subcommittee D01.35 on Solvents, Plasticizers, and Chemical Intermediates.

Current edition approved ~~Nov. 10, 1997~~; 2000. Published ~~August 1998~~; July 2000. Originally published as D 1209 – 52. Last previous edition D 1209 – 937.

- D 365 Test Methods for Soluble Nitrocellulose Base Solutions³
- D 1193 Specification for Reagent Water⁴
- D 1209 Test Method for Color of Clear Liquids (Platinum-Cobalt Scale)⁵
- E 202 Test Methods for Analysis of Ethylene Glycols and Propylene Glycols⁶
- E 346 Test Methods for Analysis of Methanol⁶

3. Significance and Use

3.1 The property of color of a solvent varies in importance with the application for which it is intended, the amount of color that can be tolerated being dependent on the color characteristics of the material in which it is used. The paint, varnish, and lacquer solvents, or diluents commercially available on today's market normally have little or no color. The presence or absence of color in such material is an indication of the degree of refinement to which the solvent has been subjected or of the cleanliness of the shipping or storage container in which it is handled, or both.

3.2 For a number of years the term "water-white" was considered sufficient as a measurement of solvent color. Several expressions for defining "water-white" gradually appeared and it became evident that a more precise color standard was needed. This was accomplished in 1952 with the adoption of Test Method D 1209 using the platinum-cobalt scale. This test method is similar to the description given in *Standard Methods for the Examination of Water and Waste Water*⁷ and is referred to by many as "APHA Color." The preparation of these platinum-cobalt color standards was originally described by A. Hazen in the *American Chemical Journal*⁸ in which he assigned the number 5 (parts per ten thousand) to his platinum-cobalt stock solution. Subsequently, in their first edition (1905) of *Standard Methods for the Examination of Water*, the American Public Health Association, using exactly the same concentration of reagents, assigned the color designation 500 (parts per million) which is the same ratio. The parts per million nomenclature is not used since color is not referred directly to a weight relationship. It is therefore recommended that the incorrect term "Hazen Color" should not be used. Also, because it refers primarily to water, the term "APHA Color" is undesirable. The recommended nomenclature for referring to the color of organic liquids is "Platinum-Cobalt Color, Test Method D 1209."

3.3 The petroleum industry uses the Saybolt colorimeter Test Method D 156 for measuring and defining the color of hydrocarbon solvents; however, this system of color measurement is not commonly employed outside of the petroleum industry. It has been reported by various sources that a Saybolt color of +25 is equivalent to 25 in the platinum-cobalt system or to colors produced by masses of potassium dichromate ranging between 4.8 and 5.6 mg dissolved in 1 L of distilled water. Because of the differences in the spectral characteristics of the several color systems being compared and the subjective manner in which the measurements are made, exact equivalencies are difficult to obtain.

4. Apparatus

4.1 *Spectrophotometer*, equipped for liquid samples and for measurements in the visible region.

NOTE 2—The spectrophotometer used must be clean and in first-class operating condition. The instrument should be calibrated in accordance with the instructions given in the Standards for Checking the Calibration of Spectrophotometers (200 to 1000 nm).⁹

4.2 *Spectrophotometer Cells*, matched having a 10-mm light path.

4.3 *Color Comparison Tubes*—Matched 100-mL, tall-form Nessler tubes, provided with ground-on, optically clear, glass caps. Tubes should be selected so that the height of the 100-mL graduation mark is 275 to 295 mm above the bottom of the tube.

4.4 *Color Comparator*—A color comparator constructed to permit visual comparison of light transmitted through tall-form, 100-mL Nessler tubes in the direction of their longitudinal axes. The comparator should be constructed so that white light is passed

² *Annual Book of ASTM Standards*, Vol 05.01.

³ *Annual Book of ASTM Standards*, Vol 06.02.

⁴ *Annual Book of ASTM Standards*, Vol 11.01.

⁵ *Annual Book of ASTM Standards*, Vol 06.04.

⁶ *Annual Book of ASTM Standards*, Vol 15.05.

⁷ *Standard Methods for the Examination of Water and Waste Water*, M. Franson, Ed., American Public Health Assoc., 14th ed., 1975, p. 65.

⁸ Hazen, A., "New Color Standard for Natural Waters," *American Chemical Journal*, Vol XIV, 1892, p. 300-310.

⁹ See NIST Letter Circular LC-1017.

TABLE 1 Absorbance Tolerance Limits For No. 500 Platinum-Cobalt Stock Solution

Wavelength, nm	Absorbance
430	0.110 to 0.120
455	0.130 to 0.145
480	0.105 to 0.120
510	0.055 to 0.065

TABLE 2 Platinum-Cobalt Color Standards

Color Standard Number	Stock Solution, mL	Color Standard Number	Stock Solution, mL
5	1	70	14
10	2	100	20
15	3	150	30
20	4	200	40
25	5	250	50
30	6	300	60
35	7	350	70
40	8	400	80
50	10	450	90
60	12	500	100 ^A

^A This is platinum-cobalt color No. 10 in Methods D 365.

TABLE 3 Platinum-Cobalt Color Standards for Very Light Colors

Color Standard Number	Stock Solution, mL	Color Standard Number	Stock Solution, mL
1	0.20	9	1.80
2	0.40	10	2.00
3	0.60	11	2.20
4	0.80	12	2.40
5	1.00	13	2.60
6	1.20	14	2.80
7	1.40	15	3.00
8	1.60		

through or reflected off a white glass plate and directed with equal intensity through the tubes, and should be shielded so that no light enters the tubes from the side.¹⁰

5. Reagents

5.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.¹¹ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Type IV of Specification D 1193.

5.3 *Cobalt Chloride* ($\text{CoCl}_2 \cdot 6\text{H}_2\text{O}$).

5.4 *Hydrochloric Acid* (sp gr 1.19) —Concentrated hydrochloric acid (HCl).

5.5 *Potassium Chloroplatinate* (K_2PtCl_6).

6. Platinum-Cobalt Reference Standards

6.1 *Platinum-Cobalt Stock Solution*— Dissolve 1.245 g of potassium chloroplatinate (K_2PtCl_6) and 1.00 g of cobalt chloride ($\text{CoCl}_2 \cdot 6\text{H}_2\text{O}$) in water. Carefully add 100 mL of hydrochloric acid (HCl, sp gr 1.19) and dilute to 1 L with water. The absorbance of the 500 platinum-cobalt stock solution in a cell having a 10-mm light path, with reagent water in a matched cell as the reference solution,¹² must fall within the limits given in Table 1.

NOTE 3—This stock solution is commercially available from reputable chemical suppliers.

6.2 *Platinum-Cobalt Standards*—From the stock solution, prepare color standards in accordance with Table 2 by diluting the required volumes to 100 mL with water in the Nessler tubes. Cap the tubes and seal the caps with shellac or a waterproof cement. When properly sealed and stored, these standards are stable for at least 1 year and do not degrade markedly for 2 years.¹³

6.2.1 For a more precise measurement of light colors below 15 platinum-cobalt, prepare color standards from the stock solution in accordance with Table 3 by diluting the required volumes to 100 mL with water in the Nessler tubes. Use a semi-microburet for measuring the required amount of stock solution.

¹⁰ The sole source of supply of the unit known to the committee at this time is Scientific Glass and Instruments, Inc., P.O. Box 6, Houston, TX 77001. If you are aware of alternative suppliers, please provide this information to ASTM Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee,¹ which you may attend.

¹¹ This test method is under the jurisdiction of ASTM Committee D-1 on Paint and Related Coatings, Materials, and Applications and is the direct responsibility of Subcommittee D01.35 on Solvents, Plasticizers, and Chemical Intermediates.

Current edition approved May 10, 2000. Published July 2000. Originally published as D 1209 – 52. Last previous edition D 1209 – 97.

¹² See the manufacturer's instruction manual for complete details for operating the spectrophotometer.

¹³ Scharf, W. W., Ferber, K. H., and White, R. G., "Stability of Platinum-Cobalt Color Standards," *Materials Research and Standards*, Vol 6, No. 6, June 1966, pp. 302–304.

7. Procedure

7.1 Introduce 100 mL of specimen into a Nessler tube, passing the specimen through a filter if it has any visible turbidity. Cap the tube, place in the comparator, and compare with the standards.

8. Report

8.1 Report as the color the number of the standard that most nearly matches the specimen. In the event that the color lies midway between two standards, report the darker of the two.

8.2 If, owing to differences in hue between the specimen and the standards, a definite match cannot be obtained, report the range over which an apparent match is obtained, and report the material as “off-hue.”

9. Precision

9.1 Color Samples with Pt-Co Color Greater than 25¹⁴:

9.1.1 These precision statements are based upon an interlaboratory study in which six platinum-cobalt standards having values of 25, 75, 170, 265, 385, and 475 were prepared in accordance with the instructions given in Section 6 of this test method and were given coded labels. These solutions were tested by one analyst in each of ten different laboratories making a single observation on one day and then repeating the observation on a second day. The analysts were requested to estimate the color to the nearest one unit for solutions below 40 platinum-cobalt, to the nearest five units for solutions between 40 and 100 platinum-cobalt and to the nearest ten units for solutions above 100 platinum-cobalt. ~~In this interlaboratory study, the within-laboratory coefficient of variation was found to be 1.8 % with 60 df, and the between-laboratories coefficient of variation was found to be 5.3 % with 54 df.~~ Based on the results of this interlaboratory study, the following criteria, calculated according to RR-D02-1007, should be used for judging the acceptability of results at the 95 % confidence level when the results are obtained under optimum conditions where the hue of the sample matches exactly the hue of the standards. Poor precision will be obtained in varying degrees as the hue of the sample departs from that of the standards.

9.1.1.1 *Repeatability*—Two results, obtained by the same analyst should be considered suspect if they differ by more than:
 $r = 0.027 (X + 92)$ platinum-cobalt units

where X is the average of the two results.

9.1.1.2 *Reproducibility*—Two results, obtained by analysts in different laboratories, should be considered suspect if they differ by more than: $R = 0.087 (X + 92)$ platinum-cobalt units

where X is the average of the two results.

9.1.1.3 Table 4 shows precision values for samples with Pt-Co Color of greater than 25.

TABLE 4 Precision Values for Greater than 25 Pt-Co Color

Color (Pt-Co units)	Repeatability, r	Reproducibility, R
25	3	10
75	5	15
165	7	22
265	10	31
385	13	41
475	16	49

9.2 Color Samples with Pt-Co Color Less than 25¹⁵:

9.2.1 The results of two interlaboratory studies were pooled to give precision values calculated according to RR-D02-1007. One study of glycols included 4 samples and 7 laboratories; the other study included 5 samples and 6 laboratories. Based on the pooled repeatability and reproducibility standard deviations, the following criteria should be used for judging, at the 95 % confidence level, the acceptability of results obtained on samples with less than 25 Pt-Co Color.

9.2.1.1 *Repeatability*— Two results, each the mean of duplicates, obtained by the same operator on different days should be considered suspect if they differ by more than two platinum-cobalt units.

9.2.1.2 *Reproducibility*— Two results, each the mean of duplicates, obtained by operators in different laboratories, should be considered suspect if they differ by more than seven platinum-cobalt units.

9.3 *Bias*—The test procedure has no bias because the value of the test result is defined only in terms of the test method.

10. Keywords

10.1 clear liquids; color; platinum-cobalt color scale

¹⁴ Supporting data are available from ASTM Headquarters. Request RR: D01-1024 and RR: D02-1007.

¹⁵ These precision statements are based on interlaboratory studies conducted by Committee E-15 on Industrial Chemicals on samples of ethylene glycol and methanol as reported in Test Methods E 202, E 346, and research reports RR: E15-28, and D01-1108.

 **D 1209 – 9700**

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).