



Standard Practice for High-Resolution Gamma-Ray Spectrometry of Water¹

This standard is issued under the fixed designation D 3649; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers the measurement of gamma-ray emitting radionuclides in water by means of gamma-ray spectrometry. It is applicable to nuclides emitting gamma-rays with energies greater than 20 keV. For typical counting systems and sample types, activity levels of about 40 Bq are easily measured and sensitivities as low as 0.4 Bq are found for many nuclides **(1)**.² Count rates in excess of 2000 counts per second should be avoided because of electronic limitations. High count rate samples can be accommodated by dilution or by increasing the sample to detector distance.

1.2 This practice can be used for either quantitative or relative determinations. In tracer work, the results may be expressed by comparison with an initial concentration of a given nuclide which is taken as 100 %. For radioassay, the results may be expressed in terms of known nuclidic standards for the radionuclides known to be present. In addition to the quantitative measurement of gamma radioactivity, gamma spectrometry can be used for the identification of specific gamma emitters in a mixture of radionuclides. General information on radioactivity and the measurement of radiation has been published **(2,3)**. Information on specific application of gamma spectrometry is also available in the literature **(4)**. See also Practice D 1066, Test Method D 1943, Practice D 3084, Practice D 3085, Practices D 3370, and Method E 181.

1.3 *This standard does not purport to address the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitation prior to use.*

2. Referenced Documents

2.1 ASTM Standards:

D 1066 Practice for Sampling Steam³

D 1129 Terminology Relating to Water³

D 2777 Practice for Determination of Precision and Bias of

Applicable Test Methods of Committee D-19 on Water³
D 3085 Practice for Measurement of Low Level Activity in Water⁴

D 3370 Practices for Sampling Water from Closed Conduits³

D 3648 Practices for Measurement of Radioactivity³

E 181 General Methods for Detector Calibration and Analysis of Radionuclides⁵

3. Terminology

3.1 *Definitions*—For definitions of terms used in this practice, refer to Terminology D 1129. For terms not defined in this practice or in Terminology D 1129, reference may be made to other published glossaries **(5)**.

4. Summary of Practice

4.1 Gamma ray spectra are measured with modular equipment consisting of a detector, an analyzer, memory, and a permanent data storage device.

4.2 Lithium-drifted germanium, Ge(Li), or high-purity germanium (HPGe) detectors, p-type or n-type, are used for the analysis of complex gamma-ray spectra because of their excellent energy resolution. These germanium systems, however, are characterized by high cost and require cooling with liquid nitrogen.

4.3 In a germanium semiconductor detector, gamma-ray photons produce electron-hole pairs. The charged pair is then collected by an applied electric field. A very stable low noise preamplifier is needed to amplify the pulses of electric charge resulting from gamma photon interactions. The output from the preamplifier is directly proportional to the energy deposited by the incident gamma-ray. These current pulses are fed into an amplifier of sufficient gain to produce voltage output pulses in the amplitude range from 0 to 10 V.

4.4 A multichannel pulse-height analyzer is used to determine the amplitude of each pulse originating in the detector, and accumulates in a memory the number of pulses in each amplitude band (or channel) in a given counting time. Computerized systems with stored programs and interface hardware can accomplish the same functions as hardwired multichannel analyzers. The primary advantages of the computerized system

¹ This test method is under the jurisdiction of ASTM Committee D-19 on Water and is the direct responsibility of Subcommittee D19.04 on Methods of Radiochemical Analysis.

Current edition approved July 10, 1998. Published March 1999. Originally published as D 3649 – 78. Last previous edition D 3649 – 98.

² The boldface numbers in parentheses refer to the list of references at the end of this test method.

³ *Annual Book of ASTM Standards*, Vol 11.01.

⁴ *Discontinued*—See 1988 *Annual Book of ASTM Standards*, Vol 11.02.

⁵ *Annual Book of ASTM Standards*, Vol 12.02.

include the capability of programming the multi-channel analyzer functions and the ability to immediately perform data reduction calculations using the spectral data stored in the computer memory or mass storage device (6). For a 0 to 2-MeV spectrum, two thousand or more data points are needed in order to fully utilize a germanium detector's excellent energy resolution.

4.5 The distribution of the amplitudes (pulse heights) of the pulses can be separated into two principal components. One of these components has a nearly Gaussian distribution and is the result of total absorption of the gamma-ray energy in the detector. This peak is normally referred to as the full-energy peak or photopeak. The other component is a continuous one lower in energy than that of the photopeak. This continuous curve is referred to as the Compton continuum and is due to interactions wherein the gamma photons lose only part of their energy to the detector. These two portions of the curve are shown in Fig. 1. Other peaks, such as escape peaks, backscattered gamma rays or X rays from shields, are often superimposed on the Compton continuum. Escape peaks will be present when gamma-rays with energies greater than 1.02 MeV are emitted from the sample (7). The positron formed in pair production is usually annihilated in the detector and one or both of the 511-keV annihilation quanta may escape from the detector without interaction. This condition will cause single or double escape peaks at energies of 0.511 or 1.022 MeV less than the photopeak energy. In the plot of pulse height versus count rate, the size and location of the photopeak on the pulse height axis is proportional to the number and energy of the incident photons, and is the basis for the quantitative and qualitative application of the spectrometer. The Compton continuum serves no useful purpose in photopeak analysis and must be subtracted when peaks are analyzed.

4.6 If the analysis is being directed and monitored by an online computer program, the analysis period may be terminated by prerequisites incorporated in the program. If the analysis is being performed with a modern multichannel

analyzer, analysis may be terminated when a preselected time or total counts in a region of interest or in a specified channel is reached. Visual inspection of a cathode-ray tube (CRT) display of accumulated data can also be used as a criterion for manually terminating the analysis on either type of data acquisition systems.

4.7 Upon completion of the analysis, the spectral data are interpreted and reduced to include activity of Bq (disintegration per second) or related units suited to the particular application. At this time the spectral data may be inspected on the CRT to identify the gamma-ray emitters present. This is accomplished by reading the channel number from the x-axis and converting to gamma-ray energy by multiplying by the appropriate keV/channel (system gain). If the system is calibrated for 1 keV per channel with channel zero representing 0 keV, the energy will be equal to the channel number. In some systems the channel number or gamma-ray energy in keV can be displayed on the CRT for any selected channel. Identification of nuclides may be aided by catalogs of gamma-ray spectra and other nuclear data tabulations (8).

4.7.1 Computer programs for data reduction have been used extensively although calculations for some applications can be performed effectively with the aid of a desk-top or pocket calculator (8). Data reduction of spectra taken with germanium spectrometry systems is usually accomplished by integration of the photopeaks above a definable background (or baseline) and subsequent activity calculations using a library which includes data such as nuclide name, half-life, gamma-ray energies and associated abundance (or branching ratios) (9).

5. Significance and Use

5.1 Gamma-ray spectrometry is of use in identifying radio-nuclides and in making quantitative measurements. Use of a semiconductor detector is necessary for high-resolution measurements.

5.2 Variation of the physical geometry of the sample and its relationship with the detector will produce both qualitative and




FIG. 1 Cesium-137 Spectrum

quantitative variations in the gamma-ray spectrum. To adequately account for these geometry effects, calibrations are designed to duplicate all conditions including source-to-detector distance, sample shape and size, and sample matrix encountered when samples are measured.

5.3 Since some spectrometry systems are calibrated at many discrete distances from the detector, a wide range of activity levels can be measured on the same detector. For high-level samples, extremely low-efficiency geometries may be used. Quantitative measurements can be made accurately and precisely when high activity level samples are placed at distances of 1 m or more from the detector.

5.4 Electronic problems, such as erroneous deadtime correction, loss of resolution, and random summing, may be avoided by keeping the gross count rate below 100 000 counts/min and also keeping the deadtime of the analyzer below 5 %. Total counting time is governed by the radioactivity of the sample, the detector to source distance and the acceptable Poisson counting uncertainty.

6. Interferences

6.1 In complex mixtures of gamma-ray emitters, the degree of interference of one nuclide in the determination of another is governed by several factors. If the gamma-ray emission rates from different radionuclides are similar, interference will occur when the photopeaks are not completely resolved and overlap. A method of predicting the gamma-ray resolution of a detector is given in the literature (10). If the nuclides are present in the mixture in unequal portions radiometrically, and if nuclides of higher gamma-ray energies are predominant, there are serious interferences with the interpretation of minor, less energetic gamma-ray photopeaks. The complexity of the analysis method is due to the resolution of these interferences and, thus, one of the main reasons for computerized systems.

6.2 Cascade summing⁶ may occur when nuclides that decay by a gamma-ray cascade are analyzed. Cobalt-60 is an example; 1172 and 1333-keV gamma rays from the same decay may enter the detector to produce a sum peak at 2505 keV and cause the loss of counts from the other two peaks. Cascade summing may be reduced by increasing the source to detector distance. Summing is more significant if a well-type detector is used.

6.3 Random summing is a function of counting rate and occurs in all measurements. The random summing rate is proportional to the total count squared and the resolving time of the detector. For most systems random summing losses can be held to less than 1 % by limiting the total counting rate to 1000 counts/s. Refer to Method E 181 for more information.

6.4 The density of the sample is another factor that can effect quantitative results. Errors from this source can be avoided by preparing the standards for calibration in solutions or other matrices with a density comparable to the sample being analyzed.

7. Apparatus

7.1 *Gamma Ray Spectrometer*, consisting of the following components:

7.1.1 *Detector Assembly*:

7.1.1.1 *Germanium Detector*—The detector shall have a volume of about 50 to 150 cm³, with a full width at one-half the peak maximum (FWHM) less than 2.2 keV at 1332 keV, certified by the manufacturer. A charge-sensitive preamplifier using low noise field effect transistors should be an integral part of the detector assembly. A convenient support shall be provided for samples of the desired form.

7.1.1.2 *Shield*—The detector assembly shall be surrounded by an external radiation shield made of massive metal, equivalent to 102 mm of lead in gamma-ray attenuation capability. It is desirable that the inner walls of the shield be at least 127 mm distant from the detector surfaces to reduce backscatter. If the shield is made of lead or a lead liner, the shield must have a graded inner shield of 1.6 mm of cadmium or tin lined with 0.4 mm of copper, to attenuate the 88-keV Pb X rays. The shield must have a door or port for inserting and removing samples.

7.1.1.3 *High Voltage Power/Bias Supply*—The bias supply required for germanium detectors usually provides a voltage up to 5000 V and 1 to 100 μ A. The power supply shall be regulated to 0.1 % with a ripple of not more than 0.01 %. Line noise caused by other equipment shall be removed with rf filters and additional regulators.

7.1.1.4 *Amplifier*—An amplifier compatible with the preamplifier and with the pulse-height analyzer shall be provided.

7.1.2 *Data Acquisition and Storage Equipment*:

7.1.2.1 *Data Acquisitions*—A multichannel pulse-height analyzer (MCA) or stand-alone analog-to-digital-converter (ADC) under software control of a separate computer, performs many functions required for gamma-ray spectrometry. An MCA or computer collects the data, provides a visual display, and outputs final results or raw data for later analysis. The four major components of an MCA are; the ADC, the memory, control, and input/output. The ADC digitizes the analog pulses from the detector amplifier. These pulses represent energy. The digital result is used by the MCA to select a memory location (channel number) which is used to store the number of events which have occurred with that energy. Simple data analysis and control of the MCA is accomplished with microprocessors. These processors control the input/output, channel summing over set regions of interest, and system energy calibration to name a few examples.

7.1.2.2 *Data Storage*—Because of the use of microprocessors modern MCAs provide a wide range of input and output (I/O) capabilities. Typically these include the ability to transfer any section of data to one or more of the following; terminal, line printer, cassette tape, floppy or hard disk, x-y plotter, and to computer interfaces by means of a serial or parallel port.

8. Sampling

8.1 Collect the sample in accordance with Practices D 3370.

8.2 Preserve the sample in a radioactively homogeneous state. A sample can be made radioactively homogeneous by the addition of a reagent in which the radionuclides or compounds

⁶ Refer to *Annual Book of ASTM Standards*, Vol 12.02 on Nuclear, Solar, and Geothermal Energy for more information on this subject.

of the radionuclides present would be soluble in large concentrations. Addition of acids, complexing agents, or stable chemically-similar carriers may be used to obtain homogeneity. Consideration of the chemical nature of the radionuclides and compounds present and the subsequent chemistry of the method shall indicate the action to be taken.

9. Test Specimens

9.1 *Containment*—Sample mounts and containers must have a convenient reproducible geometry. Considerations include commercial availability, ease of use and disposal, and the containment of radioactivity for protection of the working environment and personnel from contamination. The evaporation of liquid samples to dryness is not necessary and liquid samples up to several litres may be used. However, samples that have been evaporated to dryness for gross beta counting can also be gamma counted. Massive samples may cause significant self-absorption of low-energy gammas and degrade the higher-energy gammas. Therefore, it is important to calibrate the detector with standards of the same geometry and density. A beta absorber consisting of about 6 mm of aluminum, beryllium, or plastic may be used for samples that have a significant beta activity and high beta energies.

10. Calibration and Standardization

10.1 *Overview*—Put the instrumentation and detector into operation according to the manufacturer's instructions. Initial set-up includes all electronic adjustment to provide constant operating conditions consistent with the application and life expectancy of the calibrations. The analog-to-digital converter gain and threshold, amplifier gain and zero-level, and detector high voltage, or bias must be adjusted to yield an optimum energy calibration, usually 1 keV, or less, per channel (0.5 keV/channel is recommended). Modern commercial equipment is capable of linearity to the extent that the energy may be interpreted by the operator directly to the nearest 0.5 keV simply by reading the channel number of the highest channel in a peak and using the energy calibration data to calculate the energy of the peak. The energy calibration is usually accomplished with radioactive sources covering the entire range of interest. Subsequent efficiency calibrations and source analyses are performed with the same gain settings and the same high voltage setting. Efficiency calibrations are obtained by placing an appropriate volume of a radionuclide standard solution containing 100 to 10 000 Bq in a container and placing the container on the detector or in the well.

10.2 Procedure:

10.2.1 Preparation of Apparatus:

10.2.1.1 Follow the manufacturer's instructions, limitations, and cautions for the setup and the preliminary testing for all of the spectrometry equipment to be used in the analysis. This equipment would include, as applicable, detector, power supplies, preamplifiers, amplifiers, multichannel analyzers, and computing systems.

10.2.1.2 Place an appropriate volume of a standard or a mixed standard of radionuclides in a sealed container and place the container at a desirable and reproducible source-to-detector distance. The solution should provide about 100 counts/s in the peaks of interest and be made up of standard sources traceable

to a nationally certified laboratory. In all radionuclide measurements, the volumes, shape, physical and chemical characteristics of the samples, standards and their containers must be as identical as practicable for the most accurate results. If precipitates or residues are to be analyzed, then the standards must be evaporated on the same type of mount as the sample.

10.3 Energy Calibration:

10.3.1 The energy calibration (channel number of the multichannel analyzer versus the gamma-ray energy) of the detector system is accomplished at a fixed gain using standards containing known radionuclides. The standards should be in sealed containers and should emit at least four different gamma-ray energies covering the range of interest, usually 50 keV to 2000 keV in order to test for system linearity. Some commercially available nuclides suitable for energy calibration are the following:

	keV
lead-210	46.5
americium-241	59.5
cadmium-109	88
cerium-141	145
chromium-51	320
cesium-137	662
manganese-54	835
sodium-22	511 and 1275
yttrium-88	898 and 1836
cobalt-60	1173 and 1332
radium-226, equilibrated	186, 352, 609, 1120, and 1765

10.3.1.1 A mixed gamma-ray standard for energy and efficiency calibration is also available (see Fig. 2). This standard can be obtained in solid form in a user-supplied container. The radionuclidic purity of the standards should be verified periodically to ensure against accidental contamination or the presence of long-lived impurities by comparing the observed spectra with the spectra published in the literature (8).

10.3.2 A multichannel analyzer shall be calibrated to cover the range of interest. If the range is from 50 to 2000 keV, the gain of the system shall be adjusted until the cesium-137 photopeak, 662 keV, is about one-third full scale. Leaving the gain constant, locate at least three other photopeaks of different energies, covering the same range. Determine and record the multichannel analyzer channel number corresponding to the maximum count rate for each of the four gamma energies. Using rectangular coordinate paper, plot the gamma energy versus the channel number for each of the four gamma-ray energies. A linear relationship will be observed if the equipment is operating properly. Samples should not be analyzed if there is a nonlinear relationship. Calculate the slope and intercept of the line using a least squares calculation. If the spectrometry system is computerized, follow the appropriate manufacturer input instructions for the determination of the slope and intercept. During each day in which the spectrometry system is being used to analyze samples, the above sequence of operation shall be repeated using at least two different gamma-ray energies. If the slope and intercept are essentially unchanged, the energy calibration data remain valid. If an appreciable change in the slope or intercept is evident, the entire calibration procedure must be rerun.

10.4 Photon Detection Efficiency Calibration:

10.4.1 Accumulate an energy spectrum using sealed, calibrated radioactivity standards in a desired and reproducible




FIG. 2 Mixed Gamma-Ray Calibration Spectrum

counting geometry (see 10.2.1.2). At least 10 000 net counts (total counts minus the Compton continuum and ambient background) shall be accumulated in each full-energy gamma-ray peak of interest. Compare the live time of the count to the half-life of the radionuclide of interest. If the live time is greater than 5 % of the half-life, a correction factor must be applied for decay during the count (8).

10.4.2 Correct the radioactivity standard source gamma-ray emission rate for the decay from the time of standardization to the time at which the count rate is measured.

10.4.3 Calculate the full-energy peak efficiency, E_f , as follows:

$$E_f = N_p / N_g$$

where:

E_f = full-energy peak efficiency (counts per gamma ray emitted),

N_p = net gamma-ray count in the full-energy peak of interest (counts/s), and

N_g = gamma-ray emission rate (gamma rays/s). If the standard source is calibrated as to activity, the gamma-ray emission rate is given by:

$$N_g = A \cdot P_g$$

where:

A = number of nuclear decays per second, and

P_g = probability per nuclear decay for the gamma ray.

10.4.4 Many modern spectrometry systems are computerized and the determination of the gamma-ray efficiencies are determined automatically at the end of an appropriate counting interval. Refer to the manufacturer's instructions for specific output requirements.

10.4.5 Plot the values for the full-energy peak efficiency (as determined in 10.4.3) versus gamma-ray energy. The plot will allow the determination of efficiencies at energies for which standards are not available and to show that the algorithms used in computerized systems are providing valid efficiency calibrations. A typical plot is shown in Fig. 3.

10.4.6 Once the efficiencies have been determined, it is unnecessary to recalculate them unless there is a change in resolution, geometry, or system configuration.

11. Sample Measurements

11.1 After the spectrometer system has been set-up, the energy and efficiency calibrations performed, unknown specimens can be measured.

11.2 Following the general concepts of quantitative analytical chemistry, transfer the sample to the specimen container (9.1) and position it in the same manner as was done during system calibration (Section 10).

11.3 Measure the sample for a period of time long enough to acquire a gamma-ray spectrum which will meet the minimum acceptable counting uncertainty.

12. Calculation

12.1 Spectral data obtained with a germanium detector are only corrected for background when these peaks may alter the final results. In many experiments, the background may not affect the results but is still monitored to ensure the integrity of the system. Again, the practice presented here is not the only type but is conducive to available computational hardware and should be used to verify the validity of commercial software.

12.2 The underlying aim of this practice is to subtract the continuum or baseline from the spectral data where it underlies a photopeak of interest. For operator-directed calculations, the choice of the baseline level may be straight-forward. The simplest way, using a plot of the spectral data, is to draw a straight line, using judgement and experience, that best describes the baseline. Then the baseline data can be read directly from the plot and subtracted. "Stand-alone" computer programs have accomplished this but are not presented here.

12.3 Photopeaks lying on a sloping baseline or one with curvature will be analyzed, independent of method, with increased uncertainty. Use of data from these peaks should be limited to those cases where there is no other alternative.




FIG. 3 Efficiency Calibration Plot

Photopeaks that overlap with each other will also increase the uncertainty of the final result. In the case where use of overlapping peaks cannot be avoided, the experimenter may estimate the areas by assuming that the ratio of the peak areas is equal to the ratio of the peak heights. Computer programs separating overlapping peaks with varying degrees of success may be found in the literature (3, 9, 11).

12.4 In order to determine nuclide concentrations the photopeak areas corrected for background and interferences are divided by the count time and efficiency for the energy of the gamma ray being calculated to give gammas per second for the peak of interest. If, as is the case for some nuclides the

branching ratio is not accurately known and a direct calibration was made with the same nuclide, the branching ratio and efficiency will be one number that converts counts per second to disintegrations per second for the nuclide and photopeak of interest. If not, the gammas per second are converted to disintegrations per second by dividing the gammas per second by the gammas per disintegration, for the nuclide and photopeak of interest. The results are then corrected for sampling or decay, or both, as demanded by the application. The energies, half-lives, and gammas per disintegration for radionuclides that might be present in water samples are listed in Table 1.



TABLE 1 Nuclear Decay Data

Energy, keV	Nuclide	Half- Life ^A	Units	Gammas/ Disintegration	Energy, keV	Nuclide	Half- Life ^A	Units	Gammas/ Disintegration
46.5	Pb-210	22.3	Years	0.041	302.9	Ba-133	3981	Days	0.187
59.5	Am-241	432.2	Years	0.359	304.8	Ba-140	12.8	Days	0.034
59.5	U-237	6.75	Days	0.335	304.9	Kr-85m	4.48	Hours	0.137
63.3	Th-234	Long		0.038	315.9	Np-239	2.36	Days	0.016
80.1	Ce-144	284.3	Days	0.016	319.4	Nd-147	11.1	Days	0.020
80.1	I-131	8.04	Days	0.026	320.1	Cr-51	27.7	Days	0.099
81.0	Ba-133	3981	Days	0.329	326.0	Rh-101	3.0	Years	0.110
81.0	Xe-133	5.25	Days	0.371	328.0	Ac-228	Long		0.033
86.3	Cs-136	13.0	Days	0.063	328.8	La-140	40.2	Hours	0.200
86.5	Eu-155	4.53	Years	0.308	333.0	Au-196	6.183	Days	0.229
88.0	Cd-109	464.0	Days	0.036	334.2	Np-239	2.36	Days	0.020
91.1	Nd-147	11.0	Days	0.279	338.5	Ac-228	Long		0.123
92.4	Th-234	Long		0.027	340.6	Cs-136	13.0	Days	0.468
92.8	Th-234	Long		0.027	344.3	Eu-152	4869	Days	0.267
105.3	Eu-155	4.53	Years	0.205	351.9	Pb-214	Long		0.372
106.1	Np-239	2.36	Days	0.227	355.7	Au-196	6.183	Days	0.869
109.3	U-235	7 E8	Years	0.014	356.0	Ba-133	3981	Days	0.626
111.8	Te-132	78.2	Hours	0.019	364.5	I-131	8.04	Days	0.812
116.3	Te-132	78.2	Hours	0.019	366.4	Mo-99	66.0	Hours	0.014
121.8	Eu-152	4869	Days	0.284	380.5	Sb-125	2.77	Years	0.015
122.1	Co-57	270.9	Days	0.856	383.9	Ba-133	3981	Days	0.089
123.1	Eu-154	8.49	Years	0.405	391.7	Sn-113	115.1	Days	0.642
127.2	Rh-101	3.0	Years	0.880	402.6	Kr-87	1.272	Hours	0.496
133.5	Ce-144	284.3	Days	0.108	409.5	Ac-228	Long		0.021
136.0	Se-75	119.8	Days	0.590	411.1	Eu-152	13.5	Years	0.022
137.5	Co-57	270.9	Days	0.106	411.8	Au-198	2.696	Days	0.955
140.5	Mo-99	66.0	Hours	0.057	415.3	Rh-102	2.89	Years	0.021
140.5	Tc-99m	6.02	Hours	0.889	418.5	Rh-102	2.89	Years	0.094
143.8	U-235	7 E8	Years	0.109	420.4	Rh-102	2.89	Years	0.032
145.4	Ce-141	32.5	Days	0.484	423.8	Ba-140	12.8	Days	0.025
151.2	Kr-85m	4.48	Hours	0.751	427.9	Sb-125	2.77	Years	0.294
153.2	Cs-136	13.0	Days	0.075	432.6	La-140	40.2	Hours	0.029
162.6	Ba-140	12.8	Days	0.056	434.0	Ag-108m	127.0	Years	0.907
163.4	U-235	7 E8	Years	0.051	437.6	Ba-140	12.8	Days	0.015
163.9	Cs-123	13.0	Days	0.046	439.9	Nd-147	11.1	Days	0.012
163.9	Xe-131m	12.0	Days	0.020	446.8	Am-110m	251	Days	0.037
165.9	Ce-139	137.7	Days	0.800	463.0	Ac-228	Long		0.047
166.0	Kr-88	2.84	Hours	0.031	463.4	Sb-125	2.77	Years	0.106
176.3	Sb-125	2.77	Years	0.069	468.6	Rh-102m	206.0	Days	0.029
176.6	Cs-136	13.0	Days	0.136	475.0	Rh-102m	206.0	Days	0.460
181.1	Mo-99	66.0	Hours	0.065	475.1	Rh-102	2.89	Years	0.950
185.7	U-235	7 E8	Years	0.561	475.4	Cs-134	2.06	Years	0.015
186.2	Ra-226	1602	Years	0.033	477.6	Be-7	53.3	Days	0.103
192.3	Fe-59	44.6	Days	0.031	487.0	La-140	40.2	Hours	0.455
196.3	Kr-88	2.84	Hours	0.263	497.1	Ru-103	39.95	Days	0.895
197.9	Rh-101	3.0	Years	0.750	510.7	Tl-208	Long		0.090
205.3	U-235	7 E8	Years	0.047	511.0	Zn-65	244.0	Days	0.029
208.0	U-237	6.75	Days	0.217	511.0	Co-58	70.8	Days	0.300
228.2	Np-239	2.36	Days	0.107	511.0	Na-22	2.60	Years	1.808
228.2	Te-132	78.2	Hours	0.882	511.9	Ru-106	368.2	Days	0.206
233.2	Xe-133m	2.19	Days	0.103	514.0	Kr-85	10.72	Years	0.004
238.6	Pb-212	Long		0.446	514.0	Sr-85	64.84	Days	0.983
241.0	Ra-224	Long		0.040	529.5	Br-83	2.39	Hours	0.013
242.0	Pb-214	Long		0.075	529.9	I-131	20.8	Hours	0.873
244.7	Eu-152	13.5	Years	0.075	531.0	Nd-147	11.0	Days	0.131
248.0	Eu-154	8.6	Years	0.066	537.3	Ba-140	12.8	Days	0.244
249.8	Xe-135	9.08	Hours	0.899	554.3	Br-82	1.47	Days	0.706
256.3	Th-227	Long		0.067	556.6	Rh-102m	206	Years	0.020
264.7	Se-75	119.8	Days	0.591	563.2	Cs-134	2.06	Years	0.084
270.2	Ac-228	Long		0.038	569.3	Cs-134	2.06	Years	0.154
273.7	Cs-136	13.0	Days	0.127	569.7	Bi-207	38.0	Years	0.978
276.4	Ba-133	10.7	Years	0.073	583.2	Tl-208	Long		0.857
277.2	Tl-208	Long		0.024	591.7	Eu-154	8.6	Years	0.048
277.6	Np-239	2.36	Days	0.141	600.6	Sb-125	2.77	Years	0.178
279.2	Hg-203	46.62	Days	0.816	602.7	Sb-124	60.2	Days	0.979
279.5	Se-75	119.8	Days	0.252	604.7	Cs-134	2.06	Years	0.976
284.3	I-131	8.04	Days	0.061	606.7	Sb-125	2.77	Years	0.050
293.3	Ce-143	1.38	Days	0.420	609.3	Bi-214	Long		0.463
295.2	Pb-214	Long		0.192	610.3	Ru-103	39.95	Days	0.056
300.1	Pb-212	Long		0.034	614.4	Ag-108m	127.0	Years	0.907

TABLE 1 *Continued*

Energy, keV	Nuclide	Half- Life ^A	Units	Gammas/ Disintegration	Energy, keV	Nuclide	Half- Life ^A	Units	Gammas/ Disintegration
619.1	Br-82	1.47	Days	0.431	934.1	Bi-214	Long		0.032
620.3	Ag-110m	251	Days	0.028	937.5	Ag-110m	251.0	Days	0.344
621.8	Ru-106	368.2	Days	0.098	964.0	Eu-152	4869	Days	0.146
626.1	Rh-102m	206	Years	0.045	964.6	Ac-228	Long		0.055
628.1	Rh-102	2.89	Years	0.083	968.3	Sb-124	60.2	Days	0.018
631.3	Rh-102	2.89	Years	0.560	968.9	Ac-228	Long		0.175
635.9	Sb-125	2.77	Days	0.113	996.3	Eu-154	8.6	Years	0.103
637.0	I-131	8.04	Days	0.073	1001.0	Pa-234m	Long		0.006
645.8	Sb-124	60.2	Days	0.072	1004.8	Eu-154	8.6	Years	0.174
657.8	Ag-110m	251.0	Years	0.944	1038.6	Cs-134	2.06	Years	0.010
661.6	Cs-137	30.17	Years	0.852	1045.2	Sb-124	60.2	Days	0.797
665.6	Bi-214	Long		0.016	1046.6	Rh-102	2.89	Years	0.340
667.7	I-132	2.30	Hours	0.987	1048.1	Cs-136	13.0	Days	0.797
671.4	Sb-125	2.77	Years	0.018	1050.5	Ru-106	368.2	Days	0.017
677.6	Ag-110m	251.0	Days	0.106	1063.6	Bi-207	38.0	Years	0.740
687.0	Ag-110m	251.0	Days	0.065	1085.3	La-140	40.2	Hours	0.011
692.4	Rh-102	2.89	Years	0.016	1085.8	Eu-152	4869	Days	0.102
692.4	Eu-154	8.6	Years	0.017	1099.2	Fe-59	44.6	Days	0.565
695.2	Rh-102	2.89	Years	0.029	1103.2	Rh-102m	206	Days	0.029
696.5	Ce-144	284	Days	0.013	1103.2	Rh-102	2.89	Years	0.046
697.5	Rh-102	2.89	Years	0.440	1112.1	Eu-132	4869	Days	0.136
706.7	Ag-110m	251.0	Days	0.164	1112.8	Rh-102	2.89	Years	0.190
709.3	Sb-124	60.2	Days	0.014	1115.5	Zn-65	244.0	Days	0.507
713.8	Sb-124	60.2	Days	0.024	1120.3	Bi-214	Long		0.151
722.8	Sb-124	60.2	Days	0.113	1120.5	Sc-46	83.85	Days	1.000
723.0	Ag-108m	127.0	Years	0.915	1121.3	Ta-182	114.7	Days	0.350
723.3	Eu-154	8.49	Years	0.197	1131.5	I-135	6.61	Hours	0.225
724.2	Zr-95	65.0	Days	0.437	1136.0	I-132	2.30	Hours	0.030
727.2	Bi-212	Long		0.076	1155.3	Bi-214	Long		0.019
739.5	Mo-99	66.0	Hours	0.130	1167.9	Cs-134	2.06	Years	0.018
744.3	Ag-110m	251.0	Days	0.047	1173.2	Co-60	5.26	Years	0.999
751.9	La-140	40.2	Hours	0.043	1189.1	Ta-182	114.7	Days	0.163
755.1	Ac-228	Long		0.011	1212.8	Eu-152	4869	Days	0.014
756.7	Zr-95	65.0	Days	0.553	1216.0	As-76	1.097	Days	0.038
763.9	Ag-110m	251.0	Days	0.224	1221.4	Ta-182	114.7	Days	0.271
765.8	Nb-95	35.1	Days	0.998	1235.3	Cs-136	13.0	Days	0.198
766.8	Rh-102	2.89	Years	0.340	1238.2	Bi-214	Long		0.059
768.4	Bi-214	Long		0.050	1260.4	I-135	6.61	Hours	0.286
771.9	Ac-228	Long		0.017	1274.5	Eu-154	8.6	Years	0.355
772.6	I-132	2.30	Hours	0.762	1274.5	Na-22	950.4	Days	1.000
773.7	Te-131m	30.0	Hours	0.381	1291.6	Fe-59	44.6	Days	0.432
776.5	Br-82	1.47	Days	0.834	1293.6	Ar-41	1.83	Hours	0.992
777.9	Mo-99	66.0	Hours	0.046	1299.0	Eu-152	4869	Days	0.016
778.9	Eu-152	4869	Days	0.130	1325.5	Sb-124	60.2	Days	0.014
785.5	Bi-212	Long		0.013	1332.5	Co-60	5.26	Years	1.000
793.8	Te-131m	30.0	Hours	0.138	1368.2	Sb-124	60.2	Days	0.024
794.9	Ac-228	Long		0.049	1368.5	Na-24	15.0	Hours	1.000
795.9	Cs-134	2.06	Years	0.854	1377.8	Bi-214	Long		0.048
801.9	Cs-134	2.06	Years	0.087	1384.3	Ag-110m	251.0	Days	0.247
806.2	Bi-214	Long		0.012	1408.0	Eu-152	4869	Days	0.209
810.8	Co-58	70.8	Days	0.994	1436.7	Sb-124	60.2	Days	0.010
815.8	La-140	40.2	Days	0.235	1460.8	40K	1.28 E10	Years	0.107
818.0	Ag-110m	251.0	Days	0.073	1475.8	Ag-110m	251.0	Days	0.040
818.5	Cs-136	13.0	Days	1.000	1505.0	Ag-110m	251.0	Days	0.133
834.8	Mn-54	312.7	Days	1.000	1596.5	La-140	40.2	Hours	0.955
845.4	Kr-87	1.272	Hours	0.073	1678.0	I-135	6.61	Hours	0.095
852.2	Te-131m	30.0	Hours	0.206	1691.0	Sb-124	60.2	Days	0.488
860.6	Tl-208	Long		0.120	1764.5	Bi-214	Long		0.158
867.4	Eu-152	4869	Days	0.042	1770.2	Bi-207	38.0	Years	0.073
867.9	La-140	40.2	Days	0.055	1836.0	Y-88	106.66	Days	0.994
873.2	Eu-154	8.6	Years	0.115	2091.0	Sb-124	60.2	Hours	0.056
884.7	Ag-110m	251.0	Days	0.730	2118.9	Bi-214	Long		0.012
889.3	Sc-46	83.85	Days	1.000	2204.5	Bi-214	Long		0.049
898.0	Y-88	106.66	Days	0.950	2392.1	Kr-88	2.84	Hours	0.350
911.0	Ac-228	Long		0.277	2448.0	Bi-214	Long		0.016
919.7	La-140	40.2	Days	0.027	2614.5	Tl-208	Long		0.998
925.2	La-140	40.2	Days	0.071	2753.9	Na-22	15.0	Hours	0.998

^A Nuclides with "long" half-lives may be short-lived progeny of long-lived parents.

12.5 Results are expressed in terms of Bq/L. For the simple cases for fission or activation product radionuclides wherein there is no inherent detector contamination, calculate the nuclide concentration, $CONC$, by the following equation:

$$CONC = C_{net} / (f_o * V * BR * DC) \quad (1)$$

where:

C_{net} = net count rate (s^{-1}),
 V = test specimen volume, L,
 f_o = detector efficiency factor,
 BR = branching ration for the photon emission, and
 DC = radionuclide decay correction factor (correction for radioactive decay).

The total propagated uncertainty of the nuclide concentration σ_{CONC} is calculated as:

$$\sigma_{CONC} = CONC * [(\sigma_{C_{net}/C_{net}})^2 + (\sigma_{f_o/f_o})^2 + (\sigma_{V/V})^2 + (\sigma_{BR/BR})^2 + (\sigma_{DC/DC})^2]^{1/2} \quad (2)$$

where:

$\sigma_{C_{net}/C_{net}}$ = relative net counting rate uncertainty,
 σ_{f_o/f_o} = relative detector efficiency uncertainty,
 $\sigma_{V/V}$ = relative uncertainty in the sample volume measurement,
 $\sigma_{BR/BR}$ = relative uncertainty in the branching ratio of the gamma-ray emission, and
 $\sigma_{DC/DC}$ = relative uncertainty in the decay factor.

The net count rate C_{net} and counting uncertainty, $\sigma_{C_{net}}$, are defined as:

$$C_{net} = CR_s - (CR_B) * (S/B) = C_s/t_s - (C_B/t_s) * (S/B) \quad (3)$$

$$\sigma_{C_{net}} = [C_s/t_s^2 + (C_B/t_s^2) * (S/B)^2]^{1/2} \quad (4)$$

where:

CR_s = photopeak count rate (s^{-1}),
 CR_B = baseline background count rate (s^{-1}),
 C_s = photopeak counts,
 C_B = baseline counts corrected for the number of channels used,
 t_s = counting time of sample (s),
 S = number of channels in the photopeak, and
 B = number of channels used in the baseline subtract.

The *a priori* minimum detectable concentration (MDC), assuming an equal number of baseline and photopeak channels ($S=B$), is calculated using the equation:

$$MDC = 2.71/(t_s * k) + 4.65 * \sigma_B/k \quad (5)$$

where:

σ_B = $(CR_B/t_s)^{1/2}$, and
 k = $f_o * V * BR * DC$.

A more detailed discussion on the minimum detectable concentration concept can be found in Practices D 3648.

13. Precision and Bias

13.1 This practice is utilized for the measurement of a wide range of water sample types which may contain only a few (one to five) gamma-ray emitting nuclides or contain a large number (or a very complex mixture) of nuclides giving rise to complex spectra having interfering photon peaks. Furthermore, the concentration ranges and half-lives of individual nuclides present in a given sample may vary over several orders of

magnitude. Therefore, developing a round robin test plan in accordance with Practice D 2777 would not provide meaningful estimates of precision and bias.

13.1.1 Data published by the U.S. Department of Energy—Environmental Measurements Laboratory (EML) from their contractor Quality Assessment Program from 1982 to 1990 have been used to provide estimates of precision and bias. The U.S. Department of Energy—Environmental Measurements Laboratory contractor Quality Assessment Program provides about sixty laboratories with spiked samples twice a year. In almost all cases the data reported to EML was based upon a single measurement. Data published by this program from 1982 to 1990 are presented in Table 2.

14. Quality Control

14.1 Before this practice is utilized for the analysis of samples, a quality control or tolerance chart shall be established to ensure that the spectrometry system is operating within prescribed limits. The quality control or tolerance chart shall be established at the time the spectrometry system is calibrated.

14.2 Tolerance or statistical control charts are used to assure that an instrument or spectrometry system is operating to within prespecified limits of the initial calibration. A QC chart must be established at the time of the detector calibration. Repetitive measurements of a quality control source are taken to develop the tolerance or statistical control chart. An instrument QC check then is performed on a daily or prior to use basis, for infrequent use or for long counting times, to ensure proper operation. The result of the QC check shall be tabulated or plotted on the control or tolerance chart and evaluated according to Practice D 3648. Refer to Practice D 3648 for the preparation of a tolerance of statistical control chart.

14.3 The QC source for the detector efficiency check should contain several gamma-ray emitting radionuclides having half lives sufficiently long to eliminate significant decay corrections. The selection of the radionuclide should provide a calibration check at three widely spaced energies, preferable one for the lower energies (Am-241 – 59.5 keV), one at the intermediate energies (Cs-137 – 661.6 keV) and one at the higher energies (Co-60 – 1332.5 keV). This same source may be used to determine the daily energy calibration of the spectrometry system. Knowledge of the true disintegration rate of the radionuclides in the QC source is not essential as long as the source is decay corrected properly. Quality control sources are available commercially.

14.4 The full-width-half-maximum (FWHM) resolution of the detector should be evaluated daily or prior to instrument

TABLE 2 Estimation of Precision and Bias (See 13.1)

Nuclide	Concentration	Number of Studies	Number of Observations	Average Bias, %	Average Precision, S_p , %
	Range, Bq/L				
Mn-54	80 to 1600	13	434	+3	4
Co-57	6 to 195	6	221	+1	6
Co-60	30 to 185	12	403	-1	5
Cs-134	40 to 430	5	195	-5	9
Cs-137	65 to 200	15	485	+1	4
Ce-144	70 to 460	6	136	+7	7

use. Maintaining a constant photopeak resolution is extremely important for some computer software algorithms that use the detector resolution in the spectral unfolding routines. The gamma-ray photopeaks used for the detector efficiency check can also be used for the resolution check. Under routine operation and normal amplifier and ADC settings, the keV/channel gain may be insufficient to verify the manufacturer's detector resolution specification or slight changes to the manufacturer's specifications; however, it is good practice to perform a resolution check of the system to identify gross changes in the resolution of the detector at energies that bound the applicable energy range. The increase in resolution that can be tolerated should be determined for the application under consideration or for the spectral unfolding computer software algorithm. The measured FWHM resolution of the selected photopeaks should be trended and corrective actions taken when an intolerable condition becomes evident.

14.5 Evaluate the counting system background periodically. The background data shall be maintained in a logbook or plotted on a trend chart.

14.6 Precision and bias can be assessed in the following manner. The precision of an individual measurement can be approximated by the total propagated uncertainty (Section 13) and bias can be assessed by the analysis of NIST traceable spiked samples with known quantities of radioactivity. Bias should not be assessed using solutions derived from the same standards that were used for the detector calibration. Bias also can be evaluated by the participation in one of various commercial and government quality assurance programs that provide performance evaluation samples on a frequent basis.

15. Keywords

15.1 gamma-ray spectrometry; germanium detectors; high-resolution gamma-ray spectrometry; photon spectrometry; water

REFERENCES

- (1) Cooper, J. A., "Factors Determining the Ultimate Detection Sensitivity of Ge(Li) Gamma-Ray Spectrometers," *Nuclear Instruments and Methods*, Vol 82, 1970, pp. 273–277.
Head, J. H., "Minimum Detectable Photopeak Areas in Ge(Li) Spectra," *Nuclear Instruments and Methods*, Vol 98, 1972, pp. 419–428.
- (2) Adams, F., and Dams, R., *Applied Gamma-Ray Spectrometry*, Pergamon Press, New York, NY, 1970.
Knoll, Glen. F., *Radiation Detection and Measurement*, 2nd edition, John Wiley and Sons, Inc., New York, NY, 1989.
Dearnaley, G., *Semiconductor Counters for Nuclear Radiations*, 2nd Ed., John Wiley and Sons, Inc., New York, NY, 1966.
Panel on the Use of Lithium-Drifted Germanium Gamma-Ray Detectors for Research in Nuclear Physics, International Atomic Energy Agency, Vienna, 1966.
"Tenth Scintillation and Semiconductor Counter Symposium," *IEEE Transcript NS-13*, No. 3, 1966.
Test Procedures for Semiconductor Radiation Detectors, American National Standards Institute N42. 1-1969, IEEE 300-1969.
Mugleton, A. H. F., "Semiconductor Devices for Gamma-Ray, X-Ray, and Nuclear Radiation Detection," *Journal of Physics, E. Scientific Instruments*, Vol 5, No. 5, May 1972, pp. 390–404.
- (3) "Upgrading Environmental Radiation Data," *Health Physics Society Committee Report, HPSR-1*, 1980, Chapter 6.
- (4) Friedlander, G., Kennedy, J. W., Macias, E. S., and Miller, J. M., *Nuclear and Radiochemistry*, 3rd Edition, John Wiley and Sons, Inc., New York, NY, 1981.
Price, W. J., *Nuclear Radiation Detection*, McGraw-Hill Book Co., Inc., New York, NY, 1964.
Seigbahn, K., *Alpha-, Beta-, and Gamma-Ray Spectroscopy*, North Holland Publishing Co., Amsterdam, 1965.
Cooper, J. A., "Evaluation of Ge(Li) Compton Suppression Spectrometers for Non-Destructive Radiochemical Analysis," *BNWL-1285*, National Technical Information Service, U. S. Dept. of Commerce, 5285 Port Royal Rd., Springfield, VA 22151.
Cooper, J. A., Perkins, R. W., "A Versatile Ge(Li)-Na(Tl) Coincidence-Autocoincidence Gamma-Ray Spectrometer for Environmental and Biological Problems," *BNWL-SA-3972*, National Technical Information Service, U. S. Dept. of Commerce, 5285 Port Royal Rd., Springfield, VA 22151.
- Zimmer, W. H., "An Anticoincidence Guarded Ge(Li) Detector for Environmental Level Gamma Energy Analysis", *ARH-SA-101*, National Technical Information Service, U. S. Dept. of Commerce, 5285 Port Royal Rd., Springfield, VA 22151.
- (5) *American National Standard Glossary of Terms in Nuclear Science and Technology (ANSI N1.1)* American National Standards Institute, New York, NY
- (6) Cohn, C. E., Bennett, E. F., and Yule, T. J., "Hardware and Software for Nuclear Spectroscopy on the Varian Data Machines 622/i Computer," *ANL-7704*, National Technical Information Service, U. S. Dept. of Commerce, 5285 Port Royal Rd., Springfield, VA 22151.
Cross, J. R., Doub, J. A., and Stedman, J. M., "On-Line Data Reduction for Nuclear Analyzer," *Hewlett-Packard Journal*, Vol 22, No. 7, March 1971, pp. 2–10.
Zimmer, W. H., "A Systematic Peak Reduction Method for Semiconductor Detector Spectra," *ARH-18977*, National Technical Information Service, U. S. Dept. of Commerce, 5285 Port Royal Rd., Springfield, VA 22151.
- (7) Friedlander, G., and Kennedy, J. W., *Nuclear and Radiochemistry*, John Wiley and Sons, Inc., New York, NY, 1955, pp. 252–269.
Overman, R. T., and Clark, H. M., *Radioisotope Techniques*, McGraw-Hill Book Co., Inc., New York, NY, 1960, pp. 98–130.
Jarrett, A. A., "Statistical Methods Used in the Measurement of Radioactivity," *AECU-262*, Clearing House for Federal Scientific and Technical Information, Springfield, VA 22150.
Rider, P. R., *An Introduction to Modern Statistical Methods*, John Wiley and Sons, Inc., New York, NY, 1939.
Hoel, P. G., *Introduction to Mathematical Statistics*, 2nd Ed., John Wiley and Sons, Inc., New York, NY, 1954.
- (8) Lederer, C. M., Hollander, J. M., and Perlman, I., *Table of Isotopes*, John Wiley and Sons, Inc., New York, NY, 1967.
Marion, J. B., "Gamma-Ray Calibration Energies," *Nuclear Data*, Vol A4, 1968, pp. 301–319.
Reynolds, S. A., "Availability of Standards for Common Radionuclides," *Isotopes and Radiation Technology*, Vol 5, 1968, pp. 118–121.
Crouch, D. F., and Heath, R. L., "Routine Testing and Calibration Procedures for Multichannel Pulse Analyzers and Gamma-Ray Spectrometers," *ID0-16923*, Clearing House for Federal Scientific and

- Technical Information, US Department of Commerce, Springfield, VA, 22150.
- Heath, R. L., Black, W. W., and Cline, J. E., "Designing Semiconductor Systems for Optimum Performance," *Nucleonics*, Vol 24, No. 5, 1966, pp. 52–56.
- Test Procedures for Amplifiers and Preamplifiers for Semi-Conduction Radiation Detectors, American National Standards Institute N42.2-1969, IEEE 301-1969.
- "Calibration by Means of a Single Nuclide," *International Journal of Applied Radiation and Isotopes*, Vol 14, 1963, pp. 153–157.
- Hagee, G. R., Karches, G. J., and Goldin, A. S., "Determination of I131, Cs137, and Ba140 in Fluid Milk by Gamma Spectroscopy," *Talanta*, Vol 5, 1960, p. 36.
- Lederer, C. M., and Shirley, V. S., eds., *Table of Isotopes*, 7th ed., John Wiley and Sons, Inc., New York, NY, 1978.
- Kocher, D. C., "Nuclear Decay Data for Radionuclides Occurring in Routine Releases from Nuclear Fuel Cycle Facilities," *ORNL/NUREG/TM-102*, National Technical Information Service, U. S. Dept. of Commerce, 5285 Port Royal Rd., Springfield, VA 22151.
- Heath, R. L., "Gamma-Ray Spectrum Catalogue, Ge(Li) and Si(Li) Spectrometry," *ANCR-1000*.
- "A Handbook of Radioactivity Measurements Procedures," *Report 58* (1984), National Council on Radiation Protection and Measurements, Washington, DC, 20014.
- (9) Biggers, R. E., Bell, J. T., Long, E. C., and Russ, O. W., "Mathematical Resolution of Complex Overlapping Spectra with Non-Linear Least Squares Computer Techniques," *ORNL-3834*, National Technical Information Service, U. S. Dept. of Commerce, 5285 Port Royal Rd., Springfield, VA 22151.
- Gunnink, R., and Niday, J. B., "Computerized Quantitative Analysis by Gamma-Ray Spectrometry," *UCRL-51061*, Vol I-IV, Lawrence Livermore Laboratory, University of California, Livermore, CA 94550.
- Helmer, R. G., and Putnam, M. H., "GAUSS V, A Computer Program for the Analysis of Gamma-Ray Spectra from Ge(Li) Spectrometers," *ANCR-1043*.
- Lederer, C. M., "Computer Analysis of Spectra," *UCRL-18948*, National Technical Information Service, 5285 Port Royal Rd., Springfield, VA 22151.
- Routti, J. T., "SAMPO, A FORTRAN IV Program for Computer Analysis of Gamma Spectra from Ge(Li) Detectors, and for Other Spectra with Peaks," *UCRL-19452*, National Technical Information Service, 5285 Port Royal Rd., Springfield, VA 22151.
- Slates, R. V., "GELLI2 and SPAN2, FORTRAN Programs to Calculate Nuclide Abundances from Multichannel Gamma-Ray Spectra," *DP-1275*, National Technical Information Service, 5285 Port Royal Rd., Springfield, VA 22151.
- Ahmad, I., and Wahlgren, M., "Long-Lived Standards for the Efficiency Calibration of Ge(Li) Detector," *Nuclear Instruments and Methods*, Vol 99, 1972, pp. 333–337.
- Strauss, M. G., Lenkszus, F. R., and Eichholz, J. J., "Simple and Accurate Calibration Technique for Measuring Gamma-Ray Energies and Ge(Li) Detector Linearity," *NIM76*, 1969, pp. 285–294.
- Dwyer, P. S., *Linear Computations*, John Wiley and Sons, Inc., New York, NY, 1951.
- Overman, R. F., "Normalizing Gamma Spectra for Data Processing," *DP-751*, Clearing House for Federal Scientific and Technical Information, U.S. Department of Commerce, Springfield, VA 22150.
- Ferguson, A. J., "A Program for the Analysis of Gamma Ray Scintillation Spectra Using the Method of Least Squares," *AECL-1398*, Scientific Document Distribution Office, Atomic Energy of Canada Limited, Chalk River, Ontario, Canada, 1961.
- Strickfaden, W. B., and Kloepper, R. M., "IBM 704 Programs for Unfolding Complex Gamma Ray Spectra," *LA-2461*, Clearing House for Federal Scientific and Technical Information, U.S. Department of Commerce, Springfield, VA 22150.
- "Applications of Computers to Nuclear and Radiochemistry," *NAS-NS-3107*, Clearing House for Federal Scientific and Technical Information, U.S. Department of Commerce, Springfield, VA 22150.
- Schoenfield, E., "Alpha M-An Improved Computer Program for Determining Radioisotopes by Least Squares Resolution of the Gamma-Ray Spectra," *ORNL-3975*, National Technical Information Service, U.S. Department of Commerce, Springfield, VA 22150.
- (10) DeSoete, D., and Hoste, J., "Predicting Gamma Resolution," *Nucleonics*, Vol 20, No. 4, April 1962, pp. 72–76.
- (11) Propst, R. C., and Overman, R. F., "An Analog Computer for the Solution of Linear Simultaneous Equations," *DP-846*, Clearing House for Federal Scientific and Technical Information, U.S. Department of Commerce, Springfield, VA 22150.
- Eldridge, J. S., and Brooks, A. A., "Gamma-Ray Spectrometry Using Computer Library of Spectra," *Nucleonics* Vol 24, No. 4, 1966, pp. 54–58.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).